

Monte Carlo and analytic modeling of an Elekta Infinity linac with Agility MLC: Investigating significance of accurate model parameters for small radiation fields

4

Purpose: To explain the deviation observed between measured and Monaco calculated dose profiles for a small field (i.e. alternating open-closed MLC pattern). A Monte Carlo (MC) model of an Elekta Infinity linac with Agility MLC was created and validated against measurements. In addition, an analytic model which predicts the fluence at the isocenter plane was used to study the impact of multiple beam parameters on the accuracy of dose calculations for small fields.

8

Methods: A detailed Monte Carlo model of a 6 MV Elekta Infinity linac with Agility MLC was created in EGSnrc/BEAMnrc and validated against measurements. An analytic model using primary and secondary virtual photon sources was created and benchmarked against the MC simulations and the impact of multiple beam parameters on the accuracy of the model for a small field was investigated. Both models were used to explain discrepancies observed between measured/EGSnrc simulated and Monaco calculated dose profiles for alternating open-closed MLC leaves.

16

Results: MC simulated dose profiles (PDDs, cross- and in-line profiles, etc.) were found to be in very good agreements with measurements. The best fit for the leaf bank rotation was found to be 9 mrad to model the defocusing of Agility MLC. Moreover, a very good agreement was observed between results from the analytic model and MC simulations for a small field. Modifying the radial size of the incident electron beam in the BEAMnrc model improved the agreement between Monaco and EGSnrc calculated dose profiles by approximately 16 and 30% in the position of maxima and minima, respectively.

20

Conclusion: Accurate modeling of the full-width-half-maximum (FWHM) of the primary photon source as well as the MLC leaf design (leaf bank rotation, etc.) is essential for accurate calculations of dose delivered by small radiation fields when using virtual source or MC models of the beam.

24

28 1. INTRODUCTION

Monte Carlo (MC) techniques are accepted to be the most accurate method of dose calculation in radiotherapy and a reliable tool for modeling linear accelerators (linacs).¹ Creating a reliable dose calculation tool requires accurate and detailed knowledge of the geometry and material of the linac components as well as the characteristics of the incident electron beam through a precise benchmarking of MC model against measurements.^{1, 2}

Many researchers have studied several linac designs using MC codes to model the geometry of the treatment head and to derive beam parameters for different beam energies.³⁻¹² The methodology adopted by these groups was to create a model of the linac head based on the vendor provided information and to match depth-dose and dose-profile curve simulations against measurements to determine the initial beam parameters. In some cases, as reported by Chibani and Ma⁸, corrections to the information provided by the vendor might be required.

The sensitivity of the linac model to different parameters has been investigated by several groups.^{5, 6, 9, 11, 13-15} Sheikh-Bagheri and Rogers^{5, 6} studied beam parameters of nine megavoltage photon beams from different manufacturers (Varian, Elekta and Siemens) and concluded that MC simulations of photon beams are highly sensitive to the radial intensity and mean energy of the incident electron beam. They also reported that the accuracy of simulations is sensitive to the primary collimator opening and flattening filter material and density. Chibani and Ma^{8, 9} investigated the influence of different parameters of the incident electron beam on Varian photon beams of different energies. In addition to confirming results reported by Sheikh-Bagheri and Rogers^{5, 6}, they showed that large field sizes (e.g. 35×35 cm²) are quite sensitive to the angular divergence of the electron beam.⁹ Keall et al.¹¹ found that MC simulations are sensitive to changes in radial distribution and mean energy of the initial electron beam as well as the target density. Other groups¹³⁻¹⁵ confirmed those results and showed that accurate tuning of the incident electron beam parameters is very important to achieve the best match between MC simulations and measurements. Although Bush et al.¹⁴, investigated the impact of deviating from Gaussian intensity distribution, the optimal shape of the electron radial intensity profile was confirmed to be Gaussian. This is the shape adopted in all studies that model beam parameters in MC simulations.

52 An alternative approach of modeling treatment beams is using virtual source models (VSMs). A VSM typically
comprises of multiple virtual sources that simulate the contributions from different components of the treatment
head. These typically consist of the photons from the target, primary collimator and flattening filter as well as
electron contamination.¹⁶⁻²¹ The data for particles (e.g. position and direction) generated by each source are derived
56 from the phase space file calculated by MC simulations and scored in a specific plane. Tuning of parameters (e.g.
virtual source size, energy fluence, weight of each source) of the VSMs can be achieved by comparison against MC
simulations and/or measurements. Chabert et al.²⁰ created a virtual source model of the Elekta Synergy 6 MV photon
beam using phase space data file calculated by the PENELOPE²² Monte Carlo code and scored below the flattening
60 filter. Their VSM model included three virtual sources including a primary source (photons from the target) and
two scattered sources (photons from the primary collimator and flattening filter). They implemented their VSM in
PENELOPE and investigated the accuracy of dose calculations and portal image prediction with regards to different
binning methods to process particle information. Sikora et al.²³ showed that for field sizes smaller than $2 \times 2 \text{ cm}^2$,
64 precise modeling of the size and contribution of the primary photon source (i.e. photons from the target) is of high
importance. They showed that to achieve a good agreement between calculated and measured cross- and in-line
profiles for a $0.8 \times 0.8 \text{ cm}^2$ field, the FWHM of the primary photon source needs to be reduced by at least 30% from
its original value determined for larger field sizes. In any VSM, all calculations related to virtual sources and
68 resultant photon fluence are according to analytic and mathematical functions describing the source properties.
Besides less complexity, another advantage of using VSMs is faster calculation time compared to full MC
simulation.

The Virtual source model implemented in Monaco treatment planning system (Elekta AB, Stockholm, Sweden)
72 for Elekta linear accelerators is based on the VSM introduced by Sikora et al.¹⁹. Their model was initially created
for the Elekta Precise SLi linac and includes three virtual sources: 1) primary photon source to model photons
generated in the target; 2) secondary photon source to model photons scattered from the primary collimator,
flattening filter, anti-backscatter plate and the rest of the linac head components; 3) electron contamination source.
76 All three sources are defined to have a spatial Gaussian distribution. The primary source has a fixed radial
distribution and two other sources have energy dependent radial distributions. Particle and energy fluence for each

source are derived from appropriate phase space data stored during MC simulations. Model parameters (e.g. contribution of each source, source size) are then adjusted by comparing calculated dose in water phantom in Monaco against water measurements for an individual linac. In the Monaco beam model, the MLC as well as the jaws are included and modeled using transmission probability filters. Resulting particles from the model are finally used as input to the XVMC (X-ray Voxel Monte Carlo)²⁴ dose calculation algorithm for dose calculations within the patient.

The relationship between appropriate modeling of the MLC leaf parameters and accurate dose calculations in treatment planning systems has been investigated by several research groups²⁵⁻²⁷. Bedford et al.²⁵ verified the performance of the Agility MLC model implemented in the beam model of the Pinnacle³ (Philips Radiation Oncology System, Fitchburg, WI, USA) for calculations and delivery of VMAT plans. Kinsella et al.²⁶ and Synder et al.²⁷ used specifically designed measurements to fine tune model parameters (e.g. leaf transmission, groove width, inter-leaf leakage) of the Agility MLC model implemented in the Monaco treatment planning system.

Recent years have seen rapid improvements in the techniques of radiation therapy delivery for cancer treatment. More advanced techniques like intensity modulated radiation therapy (IMRT), volumetric modulated radiation therapy (VMAT) and stereotactic body radiation therapy (SBRT) rely on small radiation fields for high precision conformal dose delivery to a target volume while sparing organs at risk (OAR). This introduces important challenges as small fields are associated with greater uncertainty in the accuracy of beam modeling and clinical dosimetry.²⁸⁻³¹ These challenges include charged particle disequilibrium, source occlusion and choice of small detectors (e.g. small ion chambers and diodes) to reduce the effect of volume averaging of large detectors.^{28,30,31} A small radiation field is defined as one whose dimensions are comparable to or less than the lateral range of charged particles.³² Based on this criterion, for a 6 MV photon beam, field sizes equal to or less than 3×3 cm² are considered to be small.²⁸

The focus of this work is to present a detailed MC model of an Elekta Infinity linac with Agility MLC leaves. This MC model was created in EGSnrc/BEAMnrc and benchmarked against appropriate measurements. MLC model parameters were tuned using an alternating open closed field which is highly sensitive to model parameters. Comparison of dose profiles obtained from Monaco calculations and measurements/simulations for this same field

104 revealed large discrepancies. Therefore, an analytic model using multiple virtual sources was created to investigate
the impact of different beam parameters on the photon fluence at the isocenter plane. Results from the analytic
model and EGSnrc simulations are used to explain the aforementioned dose discrepancies between Monaco
calculations and measurements.

108 2. METHODS AND MATERIALS

2.A. Monte Carlo user codes

All Monte Carlo simulations were performed using EGSnrc³³ (V4-2.4.0, National Research Council of Canada,
Ottawa, ON, Canada). The BEAMnrc user code³⁴ was used to model the Elekta Infinity linear accelerator with
112 Agility MLC. Dose calculations in the water phantom were performed using the DOSXYZnrc user code³⁵.
Calculated dose from Monte Carlo simulations was converted to absolute dose using the following formulation:

$$D \text{ (cGy)} = \frac{\left(\frac{D}{\# \text{ of incident particles}}\right)_{\text{MC individual simulation}}}{\left(\frac{D}{\# \text{ of incident particles}}\right)_{\text{MC calibration simulation}}} \times \frac{1 \text{ cGy}}{\text{MU}} \times \text{MU}_{\text{del}} \quad (1)$$

where, MU_{del} is the monitor units (MU) delivered by a linear accelerator. In this formula $\frac{D}{\# \text{ of incident particles}}$
116 represents the dose scored per number of incident particles in MC simulations. The calibration simulation was
performed in water for a square field of $10 \times 10 \text{ cm}^2$ and SSD of 100 cm and dose was scored at a depth of 10 cm.

2.B. BEAMnrc model of an Elekta Infinity linear accelerator

A detailed model of an Elekta Infinity linac with Agility MLC was created using BEAMnrc. The linac model
120 was constructed for 6 MV energy, based on the technical data provided by the manufacturer and previously
published work¹⁰. A geometrical illustration of this model including the patient independent (target, primary
collimator, flattening filter, monitor ion chamber and backscatter plate) and patient dependent (160 leaves, lower
jaws) components is shown in Fig. 1.

124




FIG. 1. BEAMnrc preview of the Elekta Infinity linac model with Agility MLC showing the various component modules.

To model the multi-leaf collimator and lower jaws, the SYNCMLCE and SYNCMLCQ component modules (CMs) were used, respectively. The ‘SYNC’ versions of these component modules enable synchronization of the motion of the multi-leaf collimator, gantry and jaws in the linac model and dose calculation geometry by using a common, randomly generated MU index which lies between 0 and 1, to sample the configuration of the linac components for each particle history. A ‘SYNC’ version of the MLCQ component module was created by modifying this CM to read the MU index generated in the SYNCMLCE CM.

All measurements of output factors, percentage-depth-dose (PDD) and cross- and in-line profiles were performed in a water tank (Blue phantom², IBA Dosimetry, Schwarzenbruck, Germany) on an Elekta Infinity linac operating at a dose rate of 550 MU/min. For output factor measurements 100 MU were delivered for each field size. Measurements for small field sizes were performed using Exradin A16 ionization chamber (Standard imaging Inc., Middleton, WI, USA) as well as RFD photon diodes (IBA Dosimetry, Schwarzenbruck, Germany). For larger field sizes, CC13 ionization chamber (IBA Dosimetry, Schwarzenbruck, Germany) was used. PDD and profiles were measured using CC13 ionization chamber while build-up and penumbra regions were measured using RFD photon diodes. All measurements were performed in a water tank at an SSD equal to 100 cm. PDD curves were measured

for a $5 \times 5 \text{ cm}^2$ field size while output factors as well as cross- and in-line profiles were measured for various small and large field sizes ($2 \times 2 \text{ cm}^2$ up to $40 \times 40 \text{ cm}^2$) at 5 cm depth.

One feature of the Agility MLC is defocusing of the leaf bank (i.e. leaf bank rotation) rather than using a tongue & groove to reduce the inter-leaf leakage. In order to extract the leaf bank rotation (LBROT) value a field with alternating open-closed MLC leaf pairs was created in Monaco V.5.11.01 (Elekta AB, Stockholm, Sweden). Dose profiles for this small field size (i.e. 5 mm along the in-line/leaf bank rotation direction) are highly sensitive to LBROT variations. Two sections with 5 adjacent leaves open or closed (i.e. 2.5 cm along the in-line and 10 cm in the cross-line directions) were included in the pattern as well, to compare dose values for larger field sizes. A beam's eye view (BEV) of the described field is shown in Fig. 2. Dose profiles for this field were measured using calibrated Gafchromic film (EBT3, Ashland, Wayne, NJ, United States). Point doses were measured using an A1SL ionization chamber (Standard imaging Inc., Middleton, WI, USA) at the center of the 5-open and 5-closed leaves sections, 5.75 cm laterally from the isocenter. Dose measurements were performed at SAD of 100 cm at a depth of 5 cm in a $30 \times 30 \times 10 \text{ cm}^3$ Solid Water (RMI457 Gammex, Wisconsin, USA) phantom.




FIG. 2. BEV of the fields constructed to evaluate the LBROT value. The fields consist of a small field size (1-open leaf) to verify leaf bank rotation and one larger field size (5-open leaves) for dosimetric verification.

Source 19³⁶ was used as the particle source in BEAMnrc to define the incident electron beam. Source 19 is an elliptical beam source with Gaussian intensity distributions defined in X (cross-line) and Y (in-line) directions, with

either parallel or angular spread. The energy of the mono-energetic electron beam was determined by matching PDD curves from measurements and MC calculations for a 5×5 cm² field. Penumbrae of measured and simulated profiles in both cross- and in-line directions were matched to extract the FWHM value of the electron beam source along both directions independently. Small fields of sizes 2×2 cm² and 5×5 cm² were chosen because dose profiles of small fields are less sensitive to the mean angular divergence of the initial electron beam.⁹ The mean angular divergence of the electron beam was adjusted by comparing calculated and measured dose profiles of large fields (i.e. 30×30 cm² and 40×40 cm²). The FWHM value was then further tuned by matching measured and simulated relative output factors (ROFs).

As for the LBROT, measured dose profiles for the alternating open-closed MLC pattern were compared to simulations with varying LBROT values until the best agreement between measurements and simulations was found. The resultant leaf bank rotation causes a translation in the center of the field opening at the isocenter. The appropriate shift, calculated by following equation, was applied to the SYNCMLCE component in the beam model of the linac to account for this translation.

$$\text{shift}_{\text{MLC}} = (\text{Leaf thickness} \times \sin(\text{LBROT}))/2 \quad (2)$$

This shift was confirmed by calculating the offset between the center of the open fields of sizes 1×5, 5×5 and 15×15 cm² from MC simulations and measurements. Measurements were performed at SAD equal to 100 cm and 100 MU was delivered to the film at 5 cm depth of the solid water.

The MLC density, composition, and inter-leaf air gap were first set to the values specified by the manufacturer. These values were then adjusted to find an agreement between calculated and measured leaf transmission values utilizing the field shown in Fig. 2 and according to the explained procedure. Leaf transmission was measured using ion chamber measurements at the positions of 5-open, 5-closed leaves, respectively. The inter-leaf air gap was then tweaked until the best agreement between measurements and simulations was found for the single leaf openings in the same field. Once inter-leaf air gap value is modified, the leaf transmission changes. As a consequence, the MLC density and composition as well as the inter-leaf air gap were tweaked again to match the measured leaf transmission value. The process was then repeated until no further improvements were observed.


For all simulations, the photon cutoff energy (PCUT) and electron cutoff energy (ECUT) were set to 0.01 and 0.7 MeV, respectively, and the electron range rejection was set to 2 MeV. Bremsstrahlung cross-section enhancement was turned on and the directional bremsstrahlung photon splitting algorithm was used.

2.C. Analytic Model of the Elekta Infinity linac


An analytic model of the Elekta Infinity linac was created in Python as a simple and fast method to assist with understanding the impact of different input parameters on the photon fluence at the isocenter plane from small radiation fields. The model parameters were tuned by comparison against MC simulations using the validated BEAMnrc model of the Elekta Infinity linac as described in Section 2.B. The analytic model consists of two virtual photon sources (primary and secondary) as well as a model of the Agility MLC leaves. In order to extract parameters for the analytic model, the BEAMnrc user code was modified to score photon fluence and position as well as other photon characteristics (energy and angular distributions) at different planes of the MC model of the linac. This helped to avoid saving large phase space files that contain large number of particles to extract the needed information.

The primary photon source was used to model the bremsstrahlung photons generated in the target. In the MC linac model, photons were scored below the flattening filter, $Z = 15.9$ cm from the reference plane ($Z = 0$ cm), and back projected to the distal side of the target at $Z = 1.1$ cm. These planes are shown on Fig. 1. Only photons whose projected trajectories intersected the isocenter plane ($Z = 100$ cm) within the in-line distance of ± 10 mm from the isocenter were included in the primary source. These limits were chosen based on the small field size (5×10 mm² at isocenter) and the fact that bremsstrahlung photons coming out of the target are fairly forward-peaked. The resultant photon fluence was found to fit a Gaussian distribution with a FWHM = 1.05 mm.

The secondary photon source was created to model the scattered photons coming from the primary collimator and flattening filter and was located at the distal plane of the flattening filter at $Z = 15.9$ cm. A photon elimination criterion of ± 7 cm in the in-line distance was adopted to score photons for the secondary source. Both virtual sources are shown in Fig. 3.



(a)




(b)

FIG. 3. (a) Primary and (b) Secondary virtual photon sources used in the analytic model. The primary and secondary photon sources are placed at $Z = 1.1$ cm and $Z = 15.9$ cm from the reference plane ($Z = 0$ cm), respectively.

In order to calculate fluence profiles at the isocenter plane, a ray tracing algorithm was used to trace the trajectories of photons from both virtual sources through the MLC. The geometry of the Agility MLC leaves, including the leaf bank rotation, was extracted from MC simulations and adopted into the analytic model. The source boundaries for fluence calculation comprise the points of non-zero fluence with plus an additional 5 mm margin on either side boundaries. Attenuation of photons in the MLC leaves was modeled by exponential attenuation. Attenuation coefficient data for the leaf material composition was obtained from NIST XCOM database using the

average photon energy distribution obtained from MC simulations for each virtual source. The contribution of each virtual source at the resultant photon fluence at the isocenter plane was derived from the ratio of the number of photons from each source to the total number of photon of the fluence at the isocenter plane from MC simulations.

248 Figure 4 illustrates the ray diagram for the analytic model.



252

FIG. 4. Ray diagram illustrating the photon fluence calculation process of the analytic model.

256

The fluence at each point along the in-line position on the isocenter plane is the integral of the source. The source boundaries are shown by the photon rays tracing from the isocenter to the source plane.

2.D. Impact of analytic model parameters on the fluence at the isocenter plane

260

The impact of the analytic model parameters on the photon fluence at the isocenter plane for a single open leaf field was studied. LBROT, the size of the primary photon source, angular and energy distributions and attenuation in the MLC leaves were modified and their impact on the fluence was investigated. Moreover, since the MLC model in Monaco also includes a tongue & groove design (the Monaco beam model parameters include a tongue & groove width), the impact of parameter tongue & groove on the fluence was studied using the analytic model. Finally, the impact of exclusion of the secondary photon source was investigated. Our expectation was that the secondary photon source would have limited impact due to the fact that head scatter decreases significantly as a result of source occlusion at small field sizes³⁷. Fluence profile changes were quantified in terms of maximum photon fluence,

264

268 integral of the fluence profile and distance-to-agreement (DTA) at the penumbra region (50% of the maximum
fluence).

2.E. Comparison of Monaco and EGSnrc dose calculations

272 Dose profiles from measurements, Monaco and EGSnrc simulation were compared for the alternating open-
closed MLC pattern field described in Section 2.B. The voxelized geometry for dose calculation was defined
according to the geometry and material of the solid water phantom used for dose measurements in Section 2.B.
Voxel sizes were defined to be $1 \times 1 \times 1 \text{ mm}^3$ to match the resolution of film measurements. For dose calculations
276 using DOSXYZnrc, the linac model described in Section 2.B was used as a particle source (Source 9)³⁵ thus
eliminating the need to store a separate phase space file. A photon splitting value of 40 and other transport parameters
as described in Section 2.B were used for all DOSXYZnrc simulations. Dose was calculated at 5 cm depth of the
voxelized geometry phantom and same beam settings as explained in Section 2.B. Dose calculations were performed
280 with 2×10^9 histories to achieve a mean relative statistical uncertainty³⁸ of 0.5% over all voxels with doses greater
than 50% of the maximum dose.

Dose calculations in Monaco were performed using the XVMC²⁴ algorithm. A 1 mm dose calculation grid with
a statistical uncertainty of 1% was used for dose calculations in Monaco. Details of the beam model for the Elekta
284 Infinity linac and Agility MLC in Monaco were used to change source parameters (FWHM of the incident electron
source) and geometry of the MLC leaves (inclusion of tongue & groove with groove width of 0.4 mm) in the
BEAMnrc model of the linac.

2.F. Comparison metrics

288 For comparison purposes, point dose differences between EGSnrc and Monaco calculated dose values were
calculated. These difference comparisons were quoted as the percentage of the EGSnrc simulated dose values at the
point of comparison. Similar comparisons were performed between Monaco/EGSnrc calculations and
measurements with the difference comparisons defined as the percentage of the measured dose values at the point

292 of comparison. To compare dose profiles from EGSnrc simulations and measurements, a 1D global gamma
analysis³⁹ with various criteria (i.e. 1%/1 mm and 2%/1 mm) was utilized with measurement dose used as the
reference.

3. Results

296 3.A. BEAMnrc model of Elekta Infinity linear accelerator


Figures 5-7 show the comparison of commissioning data and MC calculations for the Elekta Infinity linac.
Excellent agreement was observed between measured and simulated PDD curves as illustrated in Fig. 5(a). All dose
points past the build-up region passed a 1%/1 mm gamma comparison. Also, over 90% of dose points from MC
300 simulations were found to be within 0.5% of measurements as shown in Fig. 5(b).

304


308

312

316



320



324


FIG. 5. (a) Comparison of the measured and calculated PDD curves for a 5×5 cm² field at SSD = 100 cm normalized at 10 cm depth, (b) Percent dose differences for calculated point doses against measurements. Error bars represent statistical uncertainty from MC simulations (0.1%).

328

The cross-line and in-line profiles (Fig. 6(a) and 6(b)) also showed good agreement between measurements and MC calculations. For cross-line profiles, all points from MC calculations passed a 2%/1 mm gamma analysis when compared against measurements. For the same criteria applied to in-line profiles, passing rates of 100% and over 95% were observed for field sizes smaller than or equal to 20×20 cm² and larger than 20×20 cm², respectively.

332

336



340




FIG. 6. Comparison of the measured and calculated (a) cross-line and (b) in-line profiles for various field sizes at 5 cm depth and 100 cm SSD.


344

Average DTA (left and right) values between MC calculated and measured data, in penumbra region (50% of the relative dose), for field sizes from 2×2 cm² to 40×40 cm² were found to be better than 0.1 mm.

Figure 7 shows a comparison of measured and MC calculated ROFs for several field sizes. The agreement was

found to be very good for all field sizes, with the largest discrepancy of less than 0.5% for the 40×40 cm² field size.

348




352

FIG. 7. Comparison of the measured and calculated relative output factors for various field sizes at 5 cm depth and 100 cm SSD.

356

Figure 8 shows the measured and simulated dose profiles for the alternating leaf field for LBROT values of 0, 6, 9 and 12 mrad. Percentage dose differences at the maxima (corresponding to 1-open leaf) and minima (corresponding to 1-closed leaf) as well as DTA values at the penumbra region (50% of the maximum dose) are shown in Table I.

360



364

368

372

FIG. 8. Comparison of dose profiles between EBT3 film measurements (solid) and MC simulations (dotted) for the field shown in Fig. 2 for LBROT values of (a) 0, (b) 6, (c) 9 and (d) 12 mrad. The best fit parameter was found to be LBROT = 9 mrad.

376


TABLE I. Mean percentage dose differences at the maxima and minima as well as average DTA (left and right) values at penumbra region from film measurements and MC simulations for different LBROT values. Uncertainties are statistical uncertainties associated with dose values at different maxima and minima from the profiles.

LBROT (mrad)	Dose difference (MC/Film) %		Average DTA (mm)
	Maxima (1-open leaf)	Minima (1-closed leaf)	
0	17.5±1.5	31.7±3.0	3.0
6	8.7±1.3	9.2±1.8	1.0
9	2.2±1.1	-5.1±1.3	0.2
12	-7.1±1.0	-14.1±1.0	1.0

From the data presented in Fig. 8 and Table I the impact of increasing leaf bank rotation on the isocenter dose profile can be seen. Also, it can be observed that the best agreement corresponds to MC simulations with LBROT value of 9 mrad. The resultant translations in the MLC leaf bank as derived from measurements and MC simulations and derived by suggested formulations were 0.42 and 0.41 mm, respectively.

Using ion chamber measurements at the position of 5-open, 5-closed leaves, the average leaf transmission was measured to be (4.3±0.1) %. From MC simulations, transmission was calculated to be (4.1±0.1) %. It was observed that a decrease of 1.1% in the density of leaves (i.e. 18.7 to 18.5 g/cm³) increases leaf transmission by 2.5%. Impact of inter-leaf air gap on leaf transmission for LBROT value of 9 mrad is shown in Fig. 9. The nominal air gap was calculated to be 0.089 mm.

396



400

FIG. 9. Variation of leaf transmission by increasing the inter-leaf air gap. All transmission values are normalized to the transmission corresponding to the nominal inter-leaf air gap.

404

From this plot we can see that as the inter-leaf air gap increases by 0.001 mm, the leaf transmission also increases by approximately 6%.

Parameters of the Elekta Infinity linac model that were derived based on the above analysis are shown in Table II.

408

TABLE II. Derived parameters of the Infinity linac model with their uncertainties.

Parameter	Value
Electron beam energy	(6.6 ± 0.1) MeV
Beam width (cross-line)	(2.1 ± 0.1) mm
Beam width (in-line)	(1.0 ± 0.1) mm
Angular divergence	(1.35 ± 0.2) deg
Leaf bank rotation angle (LBROT)	(9.0 ± 1.0) mrad
Leaf material density	18.5 g/cm ³

412 The Leaf composition (i.e. Tungsten alloy) was modified from the manufacturer provided values according to
Table III.

TABLE III. The composition of the leaf as provided by manufacturer and
adjusted in the MC beam model.

Material	Composition percentage	
	Manufacturer	MC
Tungsten (W)	95%	96%
Nickel (Ni)	3.75%	3%
Iron (Fe)	1.25%	1%

416


3.B. Analytic model of the Elekta Infinity linac

Figure 10 shows a comparison of photon fluence profiles at the isocenter plane from the analytic and Monte Carlo (BEAMnrc) linac models. Agreement of better than 1% was observed at the position of maximum fluence.

420 Also, the average DTA was found to be 0.04 mm at the penumbra region (50% of the maximum dose).

424

428



432
436
438


FIG. 10. Photon fluence at the isocenter plane from MC simulations and analytic model calculations for LBROT = 9 mrad. Fluence curves are normalized to their integral so that the integral of the resultant curve is equal to 1.

3.C. Impact of analytic model parameters on the fluence at the isocenter plane

440
442
444
446
448


The impact of modifying the analytic model parameters on the fluence at the isocenter plane, as described in Section 2.D, is illustrated in Fig. 11. From Fig. 11(a) we can see that change in the maximum fluence due to increasing leaf bank rotation follows the same trend as in Fig. 8 and Table I. Due to the source occlusion, the maximum fluence drops from approximately 40% higher to 10% lower than the fluence at the nominal LBROT (9 mrad) as leaf bank rotation increases from 0 to 12 mrad. Ignoring leaf attenuation (100% leaf transmission), as illustrated in Fig. 11(b), increases the maximum fluence by over 12% and average DTA by almost 4 times (up to 0.16 mm) compared to the scenario where attenuation by the MLC leaves is considered. It can be observed from Fig. 11(c) that when the primary source is changed from a point source to a source with radius of 2 mm the maximum fluence decreases by approximately 25% due to source occlusion. Also, the average DTA worsens once the source size deviates from the nominal value of 1 mm.

452




456

(a)



(b)

460



(c)

464

FIG. 11. Impact of changes in (a) Leaf bank rotation (normalized to nominal value of leaf bank rotation or LBROT = 9 mrad), (b) Leaf attenuation (normalized to full MLC leaf attenuation) and (c) Primary source size (normalized to nominal source size of 1 mm) on the relative fluence integral, maximum fluence and average DTA of the fluence in analytic model.

468

472

476

Modeling the energy distribution of the primary and secondary photon sources with mean photon energies of 1.6 and 0.7 MV instead of using poly-energetic sources caused a negligible fluence change of approximately 0.1%. However, a decrease of 0.2 MV in the mean energy of the primary photon source decreased the fluence integral by 1.1% due to the increase in the leaf attenuation. Using a uniform angular distribution as opposed to accurately modeling the angular distribution of the photons and its variation across the beam decreased the fluence by approximately 3% and changed the DTA to 0.05 mm. Modeling the tongue & groove in the MLC leaves caused the fluence to increase by over 2% and average DTA by 0.01 mm. Exclusion of the secondary photon source reduced the fluence by less than 1% and caused almost no change in the average DTA. This result was expected as described previously in Section 2.D.

3.D. Comparison of Monaco and EGSnrc dose calculations

Comparisons of dose profiles of the field shown in Fig. 2 for Monaco calculations against film measurements and MC simulations using EGSnrc are shown in Fig. 12(a) and Fig. 12(b), respectively.




FIG. 12. Comparison of Monaco calculations against (a) EBT3 film measurements and (b) EGSnrc simulations for the field shown in Fig. 2.

Table IV shows percentage dose differences and DTA values as similarly reported in Table I.


TABLE IV. Mean percentage dose differences at the maxima and minima as well as average DTA (left and right) values at penumbra region from Monaco calculations against film measurements and EGSnrc simulations. Uncertainties are statistical uncertainties associated to dose values at different maxima and minima from the profiles.

LBROT (mrad)	Dose difference %		Average DTA (mm)
	Maxima (1-open leaf)	Minima (1-closed leaf)	
Monaco/Film	-11.3±0.8	33.7±2.9	0.4
Monaco/EGSnrc MC	-13.2±1.0	39.5±1.5	0.5

To understand the observed differences, results of the analytic model from Sections 3.B and 3.C were used to modify the parameters in the BEAMnrc model of the linac to find a match between EGSnrc and Monaco calculations. First, the incident electron beam size (FWHM) along the in-line direction was changed from 1 to 2 mm. Next, the MLC leaf geometry was modified to include the tongue & groove as described in Section 2.E. The resultant dose profile from EGSnrc simulations is compared with Monaco calculations in Fig. 13. It can be observed

that implemented modifications in the MC model of the Elekta Infinity linac improved the agreement level between dose profiles from EGSnrc simulations and Monaco calculations.

500



504

FIG. 13. Comparison between Monaco calculations and EGSnrc simulations with modified parameters: FWHM = 2 mm in the in-line position and inclusion of tongue & groove. Dose differences were reduced to approximately 1% at the maxima and 2% at the minima.

508

Table V shows the percentage dose differences and DTA values from EGSnrc simulations with modified beam width in the in-line position as well as inclusion of tongue & groove in the MLC model.

512

TABLE V. Effect of varying model parameters in EGSnrc simulations on mean percentage differences of dose at the maxima and minima, as well as average DTA (left and right) values at penumbra regions from Monaco calculations against EGSnrc simulations. Changes were made to increase the in-line beam width to 2 mm as well as including tongue & groove to the MLC model. Uncertainties are statistical uncertainties associated with dose values at different maxima and minima from the profiles.

Parameters modified	Dose difference (Monaco/EGSnrc MC) %		Average DTA (mm)
	Maxima (1-open leaf)	Minima (1-closed leaf)	
FWHM _{In-line} = 2 mm	2.7±1.0	8.0±1.0	0.3
FWHM _{In-line} = 2mm + tongue & groove	-0.4±0.5	1.6±0.8	0.2

Benchmarking of the Monte Carlo model of a 6MV Elekta Infinity linac using the method introduced by Almgberg et al. (2012) is presented in this study. PDD curves for a $5 \times 5 \text{ cm}^2$ field as well as cross- and in-line profile measurements of different field sizes were used to derive the mean energy and radial intensity (FWHM) of the incident electron beam, respectively. Almgberg et al.¹⁰ used film measurements of the penumbra and build-up regions to take advantage of the energy independent film response. In this work similar measurements were performed using diodes combined with ion chamber to complement diode measurements and to account for the energy dependence of the diodes in large field sizes. Further adjustment of the FWHM of the radial intensity profile was performed using relative output factors. The ROFs of small fields (e.g. $2 \times 2 \text{ cm}^2$) were measured using small volume ion chamber and photon diodes. The angular distribution of the electron beam was determined from profile measurements of large field sizes. A very good agreement was found between MC calculated and measured curves for all PDD, profiles and output factor measurements. A passing rate of 100% was observed when comparing simulated PDD curves against measurements using a 1%/1 mm gamma criteria. As for cross- and in-line profiles, all simulated dose points passed a 2%/1 mm gamma comparison against measurements for field sizes smaller than or equal to $20 \times 20 \text{ cm}^2$. The passing rate for larger field sizes was better than 95%. For ROFs, worst agreement was less than 0.5% for the $40 \times 40 \text{ cm}^2$ field size.

The beam modulation system of the Infinity linac is the Agility MLC comprising 160 leaves with projected leaf width of 5 mm at the isocenter. The choice of an alternating open-closed leaves field to derive the leaf bank rotation parameter (LBROT), enabled us to apply a small field size and find the optimal value of the leaf bank rotation. An LBROT value of 9 mrad was found to give the smallest dose differences (maxima and minima) and DTA values between MC simulations and film measurements. The optimal inter-leaf air gap for this leaf geometry was found to be approximately 0.09 mm. The average measured and calculated leaf transmissions were found to agree within 5% of each other. Occasionally, appropriate corrections to vendor provided information about the material and density of different components of the linac head are required.⁸ In our study, the composition and density of the leaves, which are made of a tungsten alloy, were adjusted in the MC model for better agreement with measurements. The

fitted values of the MLC leaf composition were within 1% of the values provided by the manufacturer and density was higher by approximately 3%.

A virtual source model of our Elekta linac with Agility MLC was successfully adopted to analytically calculate the photon fluence at the isocenter plane from a small field. The model comprises of only two photon sources. The contribution of contamination electrons to the resultant fluence at the isocenter plane was neglected because it was less than 1%. The model used a simple ray tracing and exponential attenuation relationship to model the impact of the MLC leaves. Despite these approximations, our analytic model provided a simple and quick, yet reliable method, to investigate the sensitivity of the fluence to linac model parameters and to explain the disagreements between film measurements and Monaco calculations. For example, it was observed that the fluence at the isocenter plane is highly sensitive to the size of the primary photon source. Also, the integral fluence was shown to be quite sensitive to the change in the mean energy of the primary source. On the other hand, the fluence was minimally sensitive to using mean energy values for the photon sources rather than a spatial energy distribution. These results are in agreement with findings from groups who studied sensitivity of the MC model parameters to the characteristics of the incident electron beam.^{5, 6, 9, 11, 13-15} Due to the fact that the secondary photon source only represents the scattered photons, the fluence showed to have negligible sensitivity to excluding this source or changing its parameters (e.g. mean energy). However, the contribution of the secondary photon source could become more important for larger field sizes compared to the ones investigated in this study.

Regarding the impact of the leaf bank rotation, the change in the calculated fluence follows the same trend as the dose differences at the maxima in the dose profile of the alternating field as presented in Table I. Increasing the leaf bank rotation causes a decrease in the fluence due to increased occlusion of the source. Inappropriate modeling of the leaf transmission (e.g. leaf density, attenuation coefficient and leaf thickness) can also affect the fluence at the isocenter plane. However, the sensitivity of the fluence to this parameter was not found to be large since a 25% decrease in the leaf attenuation causes 4% error in the fluence. Thus, we can see that although it is important to properly model the transmission of the leaves, error of a few percent in transmission parameters is not an important source of error in the fluence calculated at the isocenter plane. This result was similar to the findings from MC simulations with modified leaf density, as explained earlier. In the analytic model, the path length of the rays

traversing the MLC leaves was defined according to their sampled trajectory. A change of less than 0.1% in the
568 fluence was observed if the rays were assumed to travel exclusively along the direction parallel to the beam axis (as
is used in Monaco beam model) rather than the accurate oblique path. This was predictable considering small
trajectory angles of the bremsstrahlung photons of the target (i.e. primary photon source) and small contribution of
the scatter photons from the primary collimator and flattening filter (i.e. secondary source).

572 Results of the analytic model calculations confirmed the findings previously shown by Sikora et al.²³ and showed
the importance of accurate modeling of the primary photon source size for the small fields. The change in the size
of the primary photon source improved the agreement between Monaco and EGSnrc calculated doses by over 10
and 30% at maxima and minima, respectively. Another important parameter in photon fluence calculations for small
576 field sizes is the modeling of the multi-leaf collimator. According to the MLC geometry parameters from the
Monaco beam model, in addition to the leaf bank rotation, a tongue & groove design with a groove width of 0.4
mm was implemented in the MLC model. To study the impact of this tongue & groove design, the leaf geometry in
our BEAMnrc model was modified using the tongue & groove width specifications in the MLC geometry file in
580 Monaco. It was found that addition of tongue & groove to the current model of MLC leaves in the BEAMnrc model
improved the agreement at both maxima and minima by about 3 and 6%, respectively, compared to the initial change
in the source size (i.e. using a source size with FWHM = 2). This is due to the insertion of groove and as a result
less occlusion of the source and higher number of photons reaching the isocenter. Similar results were observed
584 when using a leaf bank rotation of 8 mrad (no added tongue & groove) rather than 9 mrad which results in less
source occlusion as well.

From these findings we can conclude that the observed disagreements between Monaco and measured/EGSnrc
calculated dose profiles (Fig. 12(a) and Fig. 12(b)) could be associated mainly to the size of the primary source as
588 modeled in the beam model implemented in Monaco. Comparison of EGSnrc simulations using the same focal spot
size as in Monaco (i.e. FWHM = 2 mm) against Monaco calculations (Fig. 13) confirmed this conclusion as the
dose calculated by EGSnrc dropped by approximately 16% at the maxima or position of 1-open leaf. Also, since
the entire source cannot be viewed from the center of the field at the isocenter (i.e. source occlusion), the penumbra
592 region widens, which results in the reduction of the DTA between EGSnrc and Monaco calculated dose profiles and

therefore better agreement. The superposition of widened penumbrae from adjacent open leaves results in a 30% dose increase at the position of 1-closed leaf or minima of the dose profile. Overall, changing the source size from 1 to 2 mm and insertion of tongue & groove changed the result of EGSnrc simulations to be closer to the calculations obtained from Monaco.

5. CONCLUSIONS

Proper values for parameters in the Monte Carlo model of a linear accelerator head play an important role in accurate dose calculations. The same principle applies when using a virtual source model to analytically calculate the photon fluence resultant from a treatment head. Moreover, with advancements in radiation delivery techniques, using small fields has become inevitable in radiotherapy. This introduces more complexity to fine tuning model parameters of MC or analytic source models (i.e. beam and collimation parameters) due to the challenges associated with small fields.

In this paper we demonstrated the detailed MC modeling of a 6 MV Elekta Infinity linac with Agility MLC leaves that was validated against measurements. Also, we demonstrated possibility of using a simple analytic model as a quick method to study the sensitivity of different model parameters of a linac when delivering treatments with small field sizes. One important result that was studied using the analytic model and confirmed by MC simulations and measurements was the importance of adjustment of primary photon source size for small fields. Another important result obtained in this study was that modeling of the MLC leaf design (i.e. leaf bank rotation and/or tongue & groove) is essential for accurate simulation of delivered dose (maximum dose and penumbrae). Results from this study helped us explain the discrepancies observed between dose calculations obtained from Monaco treatment planning system and film measurements/EGSnrc simulations of the Elekta Infinity linac.

In conclusion, simulation of advanced techniques such as IMRT, VMAT and SBRT that comprise small fields requires a realistic MLC model as well as adjusted size of the primary photon source in the virtual source model used in treatment planning systems. Results of this study could be valuable to cancer centers that use the Elekta Infinity linac to help ensure accurate dose calculations for the above-mentioned treatment techniques.

DISCLOSURE OF CONFLICTS OF INTEREST

The authors have no relevant conflicts of interest to disclose.

- 620 ¹ I.J. Chetty, B. Curran, J.E. Cygler, *et al.*, Report of the AAPM Task Group No. 105: Issues associated with clinical implementation of Monte Carlo-based photon and electron external beam treatment planning, *Med. Phys.* **34**(12), 4818–4853 (2007).
- ² F. Verhaegen and J. Seuntjens, Monte Carlo modelling of external radiotherapy photon beams, *Phys. Med. Biol.* **48**(21), R107–R164 (2003).
624
- ³ C.M. Ma, B.A. Faddegon, D.W.O. Rogers, and T.R. Mackie, Accurate characterization of Monte Carlo calculated electron beams for radiotherapy, *Med. Phys.* **24**(3), 401–416 (1997).
- ⁴ D. Sheikh-Bagheri, D.W. Rogers, C.K. Ross, and J.P. Seuntjens, Comparison of measured and Monte Carlo
628 calculated dose distributions from the NRC linac., *Med. Phys.* **27**(10), 2256–2266 (2000).
- ⁵ D. Sheikh-Bagheri and D.W.O. Rogers, Monte Carlo calculation of nine megavoltage photon beam spectra using the BEAM code, *Med. Phys.* **29**(3), 391–402 (2002).
- ⁶ D. Sheikh-Bagheri and D.W.O. Rogers, Sensitivity of megavoltage photon beam Monte Carlo simulations to
632 electron beam and other parameters, *Med. Phys.* **29**(3), 379–390 (2002).
- ⁷ G.X. Ding, Energy spectra, angular spread, fluence profiles and dose distributions of 6 and 18 MV photon beams: results of monte carlo simulations for a varian 2100EX accelerator., *Phys. Med. Biol.* **47**(7), 1025–1046 (2002).
- ⁸ O. Chibani and C.M.C. Ma, On the discrepancies between Monte Carlo dose calculations and measurements for
636 the 18 MV varian photon beam., *Med. Phys.* **34**(4), 1206–1216 (2007).
- ⁹ O. Chibani, B. Mofteh, and C.-M.C. Ma, On Monte Carlo modeling of megavoltage photon beams: A revisited study on the sensitivity of beam parameters, *Med. Phys.* **38**(1), 188 (2011).
- ¹⁰ S.S. Almberg, J. Frengen, A. Kylling, and T. Lindmo, Monte Carlo linear accelerator simulation of megavoltage
640 photon beams: independent determination of initial beam parameters., *Med. Phys.* **39**(1), 40–7 (2012).
- ¹¹ P.J. Keall, J. V. Siebers, B. Libby, and R. Mohan, Determining the incident electron fluence for Monte Carlo-based photon treatment planning using a standard measured data set, *Med. Phys.* **30**(4), 574–582 (2003).
- ¹² L.L.W. Wang and K. Leszczynski, Estimation of the focal spot size and shape for a medical linear accelerator by

- 644 Monte Carlo simulation., *Med. Phys.* **34**(2), 485–488 (2007).
- ¹³ A. Tzedakis, J.E. Damilakis, M. Mazonakis, J. Stratakis, H. Varveris, and N. Gourtsoyiannis, Influence of initial electron beam parameters on Monte Carlo calculated absorbed dose distributions for radiotherapy photon beams., *Med. Phys.* **31**(4), 907–913 (2004).
- 648 ¹⁴ K. Bush, S. Zavgorodni, and W. Beckham, Inference of the optimal pretarget electron beam parameters in a Monte Carlo virtual linac model through simulated annealing., *Med. Phys.* **36**(6), 2309–2319 (2009).
- ¹⁵ D.L. Sawkey and B.A. Faddegon, Determination of electron energy, spectral width, and beam divergence at the exit window for clinical megavoltage x-ray beams, *Med. Phys.* **36**(3), 698–707 (2009).
- 652 ¹⁶ I. Chetty, J.J. DeMarco, and T.D. Solberg, A virtual source model for Monte Carlo modeling of arbitrary intensity distributions., *Med. Phys.* **27**(1), 166–172 (2000).
- ¹⁷ M.K. Fix, P.J. Keall, and J. V. Siebers, Photon-beam subsurface sensitivity to the initial electron-beam parameters, *Med. Phys.* **32**(4), 1164 (2005).
- 656 ¹⁸ M. Fippel, F. Haryanto, O. Dohm, F. Nüsslin, and S. Kriesen, A virtual photon energy fluence model for Monte Carlo dose calculation, *Med. Phys.* **30**(3), 301–311 (2003).
- ¹⁹ M.P. Sikora, Virtual Source Modelling of Photon Beams for Monte Carlo Based Radiation Therapy Treatment Planning, Dissertation for the degree of Philosophiae Doctor, Institute of Biomedicine, Faculty of Medicine, University of Bergen, Norway (2010).
- 660 ²⁰ I. Chabert, E. Barat, T. Dautremer, *et al.*, Development and implementation in the Monte Carlo code PENELOPE of a new virtual source model for radiotherapy photon beams and portal image calculation, *Phys. Med. Biol.* **61**(14), 5215–5252 (2016).
- 664 ²¹ A. Ahnesjö, L. Weber, A. Murman, M. Saxner, I. Thorslund, and E. Traneus, Beam modeling and verification of a photon beam multisource model, *Med. Phys.* **32**(6Part1), 1722–1737 (2005).
- ²² F. Salvat and M. Fern, P ENELOPE – A Code System for Monte Carlo Simulation of Electron and Photon Transport, OECD (November), (2003).
- 668 ²³ M. Sikora, O. Dohm, and M. Alber, A virtual photon source model of an Elekta linear accelerator with integrated mini MLC for Monte Carlo based IMRT dose calculation, *Phys. Med. Biol.* **52**(15), 4449–4463 (2007).
- ²⁴ M. Fippel, Fast Monte Carlo dose calculation for photon beams based on the VMC electron algorithm., *Med. Phys.*

26(8), 1466–1475 (1999).

- 672 ²⁵ J.L. Bedford, M.D.R. Thomas, and G. Smyth, Beam modeling and VMAT performance with the Agility 160-leaf
multileaf collimator, *J. Appl. Clin. Med. Phys.* **14**(2), 172–185 (2013).
- ²⁶ P. Kinsella, L. Shields, P. McCavana, B. McClean, and B. Langan, Determination of MLC model parameters for
Monaco using commercial diode arrays, *J. Appl. Clin. Med. Phys.* **17**(4), 37–47 (2016).
- 676 ²⁷ M. Snyder, R. Halford, C. Knill, *et al.*, Modeling the Agility MLC in the Monaco treatment planning system, *J.*
Appl. Clin. Med. Phys. **17**(3), 190–202 (2016).
- ²⁸ I.J. Das, G.X. Ding, and A. Ahnesjö, Small fields: Nonequilibrium radiation dosimetry, *Med. Phys.* **35**(1), 206–
215 (2007).
- 680 ²⁹ A.J.D. Scott, A.E. Nahum, and J.D. Fenwick, Using a Monte Carlo model to predict dosimetric properties of small
radiotherapy photon fields, *Med. Phys.* **35**(10), 4671–4684 (2008).
- ³⁰ H. Palmans, P. Andreo, S. Huq, and J. Seuntjens, Dosimetry of small static fields used in external beam
radiotherapy: An IAEA-AAPM International Code of Practice for reference and relative dose determination.
684 Technical Report Series No. 483, IAEA Trs483 (November) (2017).
- ³¹ J. Seuntjens, E. F. Lartigau, S. Cora, *et al.* ICRU Report 91, Prescribing, Recording, and Reporting of Stereotactic
Treatments with Small Photon Beams, ICRU (July) (2017).
- ³² X. A. Li, M. Soubra, J. Szanto, and L. H. Gerig, Lateral electron equilibrium and electron contamination in
688 measurements of head-scatter factors using miniphantoms and brass caps, *Med. Phys.* **22**(7), 1167–1170 (1995).
- ³³ I. Kawrakow, EGSnrc: The EGSnrc Code System : Monte Carlo Simulation of Electron and Photon Transport,
Man. - Guid. NRCC Rep. PIRS-701 (March), 1-312 (2013).
- ³⁴ D. W. O. Rogers, B. A. Faddegon, G. X. Ding, C.-M. Ma, and J. We, BEAM: A Monte Carlo code to simulate
692 radiotherapy treatment units, *Med. Phys.* **22**(5), 503-524 (1995).
- ³⁵ B. Walters, I. Kawrakow, and D.W.O. Rogers, DOSXYZnrc User's Manual, NRCC Rep. PIRS-0794, (2016).
- ³⁶ D. W. O. Rogers, B. Walters, and I. Kawrakow, BEAMnrc User's Manual, NRCC Rep. PIRS-0509, (2013).
- ³⁷ A.J.D. Scott, A.E. Nahum, and J.D. Fenwick, Monte Carlo modeling of small photon fields: Quantifying the
696 impact of focal spot size on source occlusion and output factors, and exploring miniphantom design for
small-field measurements, *Med. Phys.* **36**(7), 3132–3144 (2009).

³⁸ I.J. Chetty, M. Rosu, M.L. Kessler, *et al.*, Reporting and analyzing statistical uncertainties in Monte Carlo-based treatment planning, *Int. J. Radiat. Oncol. Biol. Phys.* **65**(4), 1249–1259 (2006).

700 ³⁹ D. a Low, W.B. Harms, S. Mutic, and J. a Purdy, A technique for the quantitative evaluation of dose distributions., *Med. Phys.* **25**(5), 656–661 (1998).