



Ultraaccurate genome sequencing and haplotyping of single human cells

Wai Keung Chu^a, Peter Edge^b, Ho Suk Lee^c, Vikas Bansal^{b,d}, Vineet Bafna^{b,1}, Xiaohua Huang^{a,1}, and Kun Zhang^{a,1}

^aDepartment of Bioengineering, University of California, San Diego, La Jolla, CA 92093; ^bDepartment of Computer Science and Engineering, University of California, San Diego, La Jolla, CA 92093; ^cDepartment of Electrical and Computer Engineering, University of California, San Diego, La Jolla, CA 92093; and ^dDepartment of Pediatrics, University of California, San Diego, La Jolla, CA 92093

Edited by George M. Church, Harvard Medical School, Boston, MA, and approved September 29, 2017 (received for review May 7, 2017)

Accurate detection of variants and long-range haplotypes in genomes of single human cells remains very challenging. Common approaches require extensive *in vitro* amplification of genomes of individual cells using DNA polymerases and high-throughput short-read DNA sequencing. These approaches have two notable drawbacks. First, polymerase replication errors could generate tens of thousands of false-positive calls per genome. Second, relatively short sequence reads contain little to no haplotype information. Here we report a method, which is dubbed SISSOR (single-stranded sequencing using microfluidic reactors), for accurate single-cell genome sequencing and haplotyping. A microfluidic processor is used to separate the Watson and Crick strands of the double-stranded chromosomal DNA in a single cell and to randomly partition megabase-size DNA strands into multiple nanoliter compartments for amplification and construction of barcoded libraries for sequencing. The separation and partitioning of large single-stranded DNA fragments of the homologous chromosome pairs allows for the independent sequencing of each of the complementary and homologous strands. This enables the assembly of long haplotypes and reduction of sequence errors by using the redundant sequence information and haplotype-based error removal. We demonstrated the ability to sequence single-cell genomes with error rates as low as 10^{-8} and average 500-kb-long DNA fragments that can be assembled into haplotype contigs with N50 greater than 7 Mb. The performance could be further improved with more uniform amplification and more accurate sequence alignment. The ability to obtain accurate genome sequences and haplotype information from single cells will enable applications of genome sequencing for diverse clinical needs.

single-cell sequencing | mutation detection | haplotyping | microfluidics

The ability to accurately identify variants in both coding regions and functional regions is essential to clinical genomics. Single nucleotide polymorphisms (SNPs) are the most common type of genetic variations. SNPs are estimated to appear in about every 100–300 bases and account for 90% of all human sequence variations (1, 2). The abundance of SNPs also provides the major source of heterogeneity for linking variants in a haplotype, where the combination of alleles occurs at multiple loci along a single chromosome. The ability to confidently identify *de novo* somatic mutations and rare germline variants on top of myriads of SNPs in single mammalian cells is challenging. Current single-cell genome sequencing approaches can manage to call millions of germline variants but also generate tens of thousands of false-positive calls that greatly outnumber the somatic mutations per genome. Besides the detection of *de novo* mutations, accurate long-range haplotyping is also useful for many clinical applications. For example, the HLA haplotype spans a ~5-Mb region in human chromosome 6. The HLA genes, which encode cell surface proteins and regulate the immune system in humans, tend to inherit as a cluster within a single haplotype. Accurate HLA haplotyping allows for better donor–patient matching for organ transplants.

Unlike targeted sequencing such as exome sequencing, whole genome sequencing (WGS) allows for the detection of all known and unknown variants. WGS using current DNA sequencing

technologies requires input DNA equivalent to tens to thousands of cells. Therefore, WGS of single cells invariably requires amplification of genomic DNA. Unfortunately, many errors are introduced in the amplification process, with error rates ranging from 1.2×10^{-5} in conventional MDA to as high as 2.1×10^{-4} in MALBAC (multiple annealing and looping-based amplification cycles) (3). In addition, current DNA sequencing technologies have substantial error rates and limited read lengths (4–6). Long-range haplotype information is very difficult or impossible to be obtained from the short sequences provided by highly parallel short-read sequencing platforms (7). The current approaches to achieve a haplotype greater than 1 Mb rely on methods to haplotype single DNA molecules. These strategies require extensive preparation in cloning (8), isolation of metaphase chromosomes (9), segregation of complementary strands in dividing daughter cells (10), or parallel partitioning of genomic DNA in water-in-oil microdroplets (11). These limitations make accurate single-cell WGS and long-range haplotyping very challenging.

Here we introduce a method called SISSOR (single-stranded sequencing using microfluidic reactors), in which double-stranded chromosomal DNA molecules from single cells are

Significance

Accurate sequencing and haplotyping of diploid genomes of single cells are intrinsically difficult due to the small amount of starting materials and limited read lengths of current DNA sequencing methods. In SISSOR (single-stranded sequencing using microfluidic reactors), we aim to improve sequencing accuracy and haplotype assembly by taking advantage of the redundant complementary sequence information in the double-stranded DNA and by partitioning megabase-size single-stranded DNA fragments from the homologous chromosome pairs into multiple compartments for amplification by MDA (multiple displacement amplification) and subsequent sequencing using short-read DNA sequencing platforms. We report the demonstration of the most accurate single-cell genome sequencing to date with data from three single human cells. Our approach can simultaneously provide higher accuracy and longer haplotypes than existing approaches.

Author contributions: X.H. and K.Z. designed research; W.K.C. and H.S.L. performed research; V. Bansal and V. Bafna contributed new analytic tools; W.K.C., P.E., H.S.L., V. Bansal, V. Bafna, X.H., and K.Z. analyzed data; and W.K.C., P.E., V. Bafna, X.H., and K.Z. wrote the paper.

Conflict of interest statement: X.H. and K.Z. are listed as inventors for a patent application related to the method disclosed in this manuscript. K.Z. is a cofounder and equity holder of Singlera Genomics Inc. V. Bafna is a cofounder, has an equity interest, and receives income from Digital Proteomics, LLC. The terms of this arrangement have been reviewed and approved by the University of California, San Diego in accordance with its conflict of interest policies. Digital Proteomics, LLC was not involved in the research presented here.

This article is a PNAS Direct Submission.

This is an open access article distributed under the PNAS license.

Data deposition: The data reported in this paper have been deposited in the NCBI Sequence Read Archive, <https://www.ncbi.nlm.nih.gov/sra> (accession no. SRP106579).

See Commentary on page 12362.

¹To whom correspondence may be addressed. Email: vbafna@cs.ucsd.edu, x2huang@ucsd.edu, or kzhang@eng.ucsd.edu.

This article contains supporting information online at www.pnas.org/lookup/suppl/doi:10.1073/pnas.1707609114/-DCSupplemental.

separated and megabase-size fragments are stochastically partitioned into multiple nanoliter compartments for enzymatic amplification in a microfluidic device. The random partitioning allows for independent amplification and sequencing of the homologous chromosomal DNA fragments for long-range haplotype assembly and error correction by comparing the phased complementary strands. As a proof of concept, we amplified and sequenced three single cells from a human fibroblast cell line (PGP1f) whose genome has been sequenced extensively using other approaches.

Results

Partitioning and Amplification of Separated Single-Stranded Chromosomal DNA Fragments. We implemented the SISSOR concept using an integrated microfluidic processor. The device and the overall procedure are illustrated in Fig. 1 (more detail in *SI Appendix, Fig. S1*). The microfluidic device consists of four modules: single-cell capture, cell lysis and strand separation, partitioning, and amplification modules. A single cell is captured from a cell suspension. The cell is lysed and double-stranded chromosomal DNA molecules are separated using an alkaline solution. The separated DNA fragments are randomly distributed using a rotary pump and partitioned into 24 identical chambers. Each partition is then neutralized and pushed down into the amplification module and amplified by MDA. Amplified products in each chamber are retrieved, converted into barcoded sequencing libraries, and sequenced using Illumina short-read sequencing-by-synthesis.

We experimented with various procedures and designs of the processor to optimize SISSOR. We found that cell lysis, denaturation of the long double-stranded chromosomal DNA molecules, and the distribution of the ssDNA fragments were more effective with a higher KOH concentration (up to 400 mM) and pumping speeds of the rotary mixer. To prevent DNA damage at high pH, the process was limited to 10 min at 20 °C. We also found that MDA chambers with a lower surface-to-volume ratio and pre-coating of the polydimethylsiloxane (PDMS) surface with BSA improved the amplification, perhaps by reducing the inhibition of the DNA polymerase and the nonspecific binding of molecules. MDA chambers with a volume of ~20 nL provided sufficient amplification with less bias by limiting the available nucleotides (dNTPs). Further reducing the MDA reaction volume decreased the genome coverage due to the lower than sufficient input of amplicons for downstream ligation of sequencing adapters (*SI Appendix, Fig. S2B and Table S1*). Similar genome coverage was observed with the use of

longer random oligo primers. Without sufficient mixing during the denaturation step, ssDNA fragments were not separated to the partition chambers (*SI Appendix, Fig. S2C*).

Genome Coverage. We amplified the genomes of three single cells from the human PGP1 fibroblast cell line using the optimized devices and procedures. The amplified genomic DNA collected from each of the individual 24 chambers was converted into barcoded sequencing libraries. The barcoded libraries from each cell were combined and sequenced using standard 100-bp paired-end Illumina sequencing. Sequencing reads from the individual chambers were identified using the barcodes and mapped to the reference human genome hs37d5 (GRCh37/b37 + decoy sequences) using the default setting of BWA-MEM with Burrows–Wheeler Aligner (BWA) (*SI Appendix, Fig. S3A*). Deep sequencing data from the three single cells yielded a combined 558 Gb, with 92.6–98.8% mappable reads. We obtained an average of 65-fold sequencing depth and $63.8 \pm 9.8\%$ genome coverage per base per cell. The combined sequence reads from three cells cover 94.9% of the entire genome (*SI Appendix, Table S2*). The high mapping rate was perhaps the result of the high-fidelity Phi-29 polymerase used in MDA and reduced contamination in small reaction volumes of the microfluidic devices. BWA-MEM also identified and removed chimeric reads commonly introduced by MDA. Since the combined genome coverage from three cells is ~30% higher than any individual cell, we suspect that some fragments were lost during strand separation and partitioning within the device and did not reach the amplification chambers.

Determination of SISSOR Fragments. Megabase-sized DNA fragments were visualized by mapping all sequencing reads to the reference genome sequence (*SI Appendix, Fig. S4*). The individual DNA fragments appeared as dense blocks with overlapping unique reads. Genome regions that have dense mapped coverage with reads from two or more chambers indicate that the four complementary strands of the homologous chromosome pairs were separated, partitioned, and amplified in different chambers (Fig. 2). The reads from each chamber that are mapped to sporadic regions of the genome with very low density coverage are very likely the consequence of misalignment. Thus, they were removed in the segmentation of SISSOR fragments using the Hidden Markov model (HMM) (*SI Appendix, SI Methods*).

We determined the boundaries of the large SISSOR fragments by joining the aligned sequencing reads using HMM, based on read depth and proximity in the localized genomic regions. We




Fig. 1. An overview of the experimental process of SISSOR technology. A single cell in suspension was identified by imaging and captured. The cell was lysed, and chromosomal DNA molecules were separated into single-stranded form using ALS. The single-stranded DNA molecules were randomly distributed and partitioned in 24 chambers. Each partition was pushed into an air-filled MDA chamber using a neutralization buffer, followed by an MDA reaction solution. MDA reaction was carried out by heating the entire device at 30 °C overnight. The amplified product in each individual chamber was collected out of the device and processed into the barcoded sequencing library.




Fig. 2. Haplotyping of single-stranded DNA fragments using sequencing reads from a single-cell genome amplified using a SISSOR device with 24 chambers. Large subhaploid SISSOR fragments were first computed per chamber and then phased into haplotype 1 and haplotype 2 with HapCUT2 (13). SISSOR fragments could be visualized by mapping the sequencing reads to a reference genome. Some fragments were not phased due to either the lack of heterozygous SNPs or the presence of mixed sequences from two or more strands.

counted the number of reads in each good bin defined by the 50K variable bin method (12) and calculated the average number of reads if these reads were randomly distributed in all 50K bins (*SI Appendix, Figs. S3B and S5*). The resulting fragment boundaries as determined by the start and end positions of continuous bins in HMM were highly consistent in the range of 1–5× average reads per bin. These boundaries closely resembled the subhaploid DNA fragments because of the high ratio of reads per bin concentrated in a small genomic region rather than distributed randomly in the entire genome. Copy number variations could potentially be detected based on the significant deviation of SISSOR fragment counts within a genomic interval, although this remains to be further established. About 11.8% of mapped locations were removed in HMM by choosing 5× average reads per bin (*SI Appendix, Table S3*). The N50 fragment size exceeded 1 Mb, with the largest contig at 9 Mb (*SI Appendix, Table S4*). The mean DNA fragment length before haplotype assembly is ~500 kb, which is 5–10-fold longer than what has been achieved using dilution methods. We estimated that each single-stranded DNA fragment was amplified about 60,000 times.

Variant Calling. The unique design of SISSOR enables more accurate single-cell variant calls in two ways: First, variants that are observed in multiple chambers can be called more confidently than variants observed only once (multiple chamber allele occurrence). Second, variant calls that match between strands from the same haplotype are of especially high confidence (same-haplotype strand matching).

To leverage multiple chamber allele occurrence, we developed a variant calling algorithm, which closely models the SISSOR workflow to make consensus allele calls for every genomic position using sequence information from all chambers (*SI Appendix, SI Methods and Fig. S3 C and D*). Briefly, the algorithm

models all of the possibilities that single DNA strands from a diploid genome could be distributed to and amplified in the chambers and accounts for possible error sources such as MDA. The algorithm assigns higher confidence to alleles that are observed multiple times and fit well into the diploid model. We gauged the accuracy of our algorithm at different confidence thresholds by comparing it to a reference genome sequence for PGP1 (*SI Appendix, SI Methods and Table S5*). At the most lenient threshold, 1.7 million SNVs were called with a false-positive rate of 5×10^{-5} . At a moderate threshold, 613,669 SNVs were called with a false-positive rate of 1×10^{-6} . At the strictest threshold, 177,096 SNVs were called with a false-positive rate of 1×10^{-7} .

Even greater accuracy can be achieved by leveraging same-haplotype strand matching, an approach that requires separating fragments into different haplotypes. To perform haplotype assembly, we extended our variant calling model to call the most likely allele in every chamber (at a lenient threshold) and generate subhaploid fragment sequences (*SI Appendix, SI Methods*). In the following sections, we describe haplotype assembly and validation of variant calls by same-haplotype strand matching to achieve maximum accuracy using the SISSOR technology.

Whole Genome Haplotyping. Haplotype assemblies were constructed by phasing heterozygous SNPs in subhaploid SISSOR fragments. A list of heterozygous SNPs, obtained from 60× coverage Illumina WGS data of PGP1 fibroblast cells (under ENCODE project “ENCSR674PQI”), was used to phase the 1.2 million SNPs in SISSOR fragments. We applied these SNPs to a haplotyping algorithm, HapCUT2 (13), and compared the assembly to the PGP1 haplotype created using subhaploid pools of BAC clones (8). Two types of errors may occur in an assembled haplotype. First, a switch error was defined as two or more SNPs in a row flipped. Second, a mismatch error was defined as a heterozygous SNP whose phase was incorrectly inferred. If a higher switch and mismatch error rate (1.6%) could be tolerated in an application, a large N50 haplotype length (>15 Mb) was directly produced by HMM-derived SISSOR fragments. We anticipate that genome resolution can be augmented by mapping high-quality short sequencing reads to the long haplotype scaffold. Similarly, long-range chromosome-length haplotype scaffolds have been created with the Strand-seq approach, which required BrdU incorporation in dividing cells (10) and hence was not applicable to nondividing cells or archived tissues. Combining the heterozygous variants in short WGS reads (~250 bp) to long haplotypes was shown to improve the phased coverage.

We further processed and refined the raw SISSOR fragments to address the case where two overlapping homologous DNA fragments may appear in the same chamber (*SI Appendix, Fig. S3D*). Long SISSOR fragments were split where the phase of two SNPs in a row are flipped with respect to fragments from other chambers. We removed the fragments with clusters of low-quality variant calls and then reassembled these processed fragments with HapCUT2. Splitting longer fragments with detectable switch errors and poor variant calls from mixed homologous reads at the unique genomic position reduced the overall haplotyping errors. Four-strand coverage of processed fragments reduced more than 17% of the original size, but the phasable whole SISSOR fragments increased from 70–80% to about 93% in all three cells.

Although the lengths of processed SISSOR fragments were reduced, HapCUT2 assembly of overlapping fragments still creates a long haplotype contig with an N50 ~ 7 Mb and with a >90% span of the human genome (Table 1). In comparison with the BAC haplotype, which has an N50 ~ 2.6 Mb for the PGP1 cell line, the 1.2 million heterozygous SNPs called and phased in our SISSOR libraries have a concordance rate of 99.3% in our assembled haplotype blocks. The achieved haplotype

Table 1. Summary of haplotyping performance

Metrics	Values
No. of haplotype blocks	1,960
Haplotype block average length, Mb	1.4
N50 haplotype contig length, Mb	7.1
Largest haplotype contig length, Mb	28.3
Total genomic span, Gb	2.63
No. of phased heterozygous SNPs	1,248,150
Switch discordance rate	0.41%
Mismatch discordance rate	0.31%

concordance in SISSOR is comparable to the 99.1% and 99.4% accuracy rates obtained using fosmid clone and PacBio 44× coverage SMRT data, respectively, for the NA12878 genome using HapCUT2 (13).

Error Rate of MDA, Library Preparation, and Sequencing. To quantify the error rate, each SISSOR fragment was assigned to a haplotype by matching two or more heterozygous SNPs with over 80% accuracy in the assembled haplotype blocks (Fig. 2 and *SI Appendix, Fig. S3E*), and the number of mismatched bases in each pair of the haplotype-matched fragments was counted. The rate of mismatched bases between phased fragments with the same haplotypes obtained from between any cells is about 1.7×10^{-5} (*SI Appendix, Table S6*). The errors are very likely introduced by MDA, PCR-based library construction, and sequencing. The error rate is found to be consistent with that of other MDA-based sequencing methods (3, 14). The unphased fragments have fewer heterozygous SNPs due to their limited length, low coverage, or being located within regions with low sequence complexity. The unphased fragments and reads outside the fragment boundaries were removed from downstream analysis.

Single-cell MDA is susceptible to trace amounts of DNA contamination in reagents and reaction vessels. The processing and amplification performed in the close environment of the microfluidic device with minimal flow paths and small volume of the microreactors minimizes contamination. This is reflected in the high mapping rate of sequencing reads and low mismatch rate we obtained (*SI Appendix, Tables S2 and S6*). Mutagenic damages associated with DNA oxidation could result in mutations, which are usually prevalent in G to T mutations (15). We minimized oxidative DNA damage by including DTT as a reducing agent in the denaturation solution. For the 3,696 mismatches between the two complementary strands with a total ~179 Mb (*SI Appendix, Fig. S6*), the rate of G to T mutation is only 2%, while the ratio of transition to transversion mutations is 3.45. This indicates that oxidative damage was minimal or negligible in the SISSOR device.

Error Rate of Base Consensus in Phased Single-Stranded DNA Fragments. SISSOR technology can improve single-cell variant calling by matching same-haplotype strands, since variant calls that match between two complementary strands of DNA are of especially high confidence. However, it is not feasible to directly measure the accuracy of this approach for a single cell, since there is no reference genome for an individual cell with its de novo mutations. Instead, we provide an indirect estimate of the maximum possible error rate of same-haplotype strand matching in SISSOR technology, by noting that same-haplotype matching variant calls between strands in different cells do not carry cell-specific mutations (Fig. 3). We sampled haplotype-matched allele calls at the same genomic position using sequencing data in two chambers from different cells (cross-cell) and compared the matching consensus to the PGP1 reference. MDA errors, single cell de novo variants, and haplotyping errors were removed by only considering the

position with identical calls in two distinct cells (Fig. 3B, position 1–2). We expect that any unvalidated calls are either true errors or true de novo SNVs found in the shared lineage of the individual cells.

Phased haplotype fragments were randomly paired once between two cells at each unique haploid position (*SI Appendix, Fig. S3F*). Unique cross-cell positions of 351 Mb co-occurred with the consensus of Complete Genomics (CGI) (16) and Illumina WGS (17) reference of PGP1 (*SI Appendix, Table S7*). Only 19 matching cross-cell calls in SISSOR were discordant to the PGP1 reference (Fig. 4). Of these calls, 10 variants were found in the BAC reference and validated as true variants. Another five variants appeared in a third SISSOR chamber independent of the two haplotype-paired strands, indicating that these cases were not double errors in the phased strand consensus (Fig. 3B, position 5). Multiple appearance of the same variant confirmed that this variant either was previously undetected or only existed as de novo mutation shared by this cell lineage. Four variants remained unaccounted for. There is not sufficient information to determine whether the variants are genuine de novo SNVs or are due to errors introduced by the SISSOR procedures. This bounds the overall sequencing error rate of the SISSOR technology below 1×10^{-8} (four possible errors in 351 Mb). Further investigation showed that one of the four variants was supported by the sequence from another chamber with a read depth of 4, which was lower than the threshold depth of 5. Another allele had PGP1 reference calls in




Fig. 3. Error rate analysis of base consensus in phased SISSOR fragments. (A) Base sequences in single-stranded DNA fragments were constructed by variant calling of the mapped MDA products in each individual chamber, and the complementary strands were identified by comparing the haplotypes of the single-stranded fragments from different chambers. (B) Matching variant calls in the contigs from the same haplotype between two cells (cross-cell), representing the PGP1-specific sequence, were validated by the PGP1/WGS reference. Common MDA and library preparation error was defined by the mismatches of variant calls between two matching phased haplotypes within the same cell (position 1). Single-cell de novo mutation was defined by matching variant calls between two matching phased haplotypes, together with a matching variant call from at least one chamber in the other cell to the PGP1 reference (position 2). The rates of single-chamber MDA-based sequencing error (10^{-3}) and single-cell de novo mutation (10^{-7}) were calculated for SISSOR. Cross-cell consensus, where de novo variants were removed, was defined by the matching variant calls between phased haplotypes in two different cells (position 3–7). The mismatch consensus to the PGP1 reference call (position 5) represented the discordance rate for SISSOR technology (10^{-8}).




Fig. 4. Differences between allele calls in PGP1 reference and SISSOR consensus. The consensus of CGI and Illumina WGS was used as the PGP1 reference. Positions lacking coverage in both PGP1 reference and SISSOR consensus were discarded.

the opposite haplotypes from the other chambers, which suggested the correct mapping and haplotyping at this location and further strengthened the possibility of these variants as true de novo calls.

Direct Measurement of de Novo Variants in a Single Cell. We expect sequencing accuracy equivalent to the strand–strand consensus in the same cell, assuming error rate was identical in all SISSOR libraries. Number of matched SNVs within a single cell was presumably higher than cross-cell matches because of the true de novo variants in a single cell. In comparison with bulk sequencing, individual mutations from each single cell are masked and undetected by the consensus of many other cells. In contrast, the consensus of phased SISSOR fragments from each single cell detected a total of 68 possible de novo SNVs (*SI Appendix, Table S8*). None of these were found in the other PGP1 libraries prepared using induced pluripotent stem cell and fibroblast (16), where their combined consensus was about 0.1% different from lymphocyte only. Correct PGP1 variant calls were found in 28 (~41%) SNVs, such as position 2 of cell 1 in Fig. 3*B*, where novel calls were discovered only in cell 2. Seven (~10%) base calls were found to have the correct PGP1 reference base called in the opposite haplotype within the same cell. Seventeen (~25%) were found to have the correct PGP1 reference base called in both the opposite haplotype and the other cell. Many of these variants appear to be real since correct PGP1 reference base calls suggest correct mapping at the identical positions. The remaining 16 SNVs lacking coverage from the other cells are potential false-positive errors. However, all but two unsupported SNVs were presumably undetected de novo variants because of the lower error in SISSOR technology.

Direct Resolution of the HLA Haplotypes. Haplotype of the HLA genes has been determined using SISSOR. The highly polymorphic HLA loci are distributed within a ~5-Mb region on chromosome 6 (28.5–33.5 M) and are important to the immune system. We obtained a single SISSOR haplotype block of ~18.2 Mb spanning the entire HLA region. This allows for the phasing of the four classical HLA genes. We selected the sorted BAM files from individual SISSOR fragments that correspond to the HLA haplotype and further determined the best genotypes of HLA genes via the tool bwakit (under github “lh3/bwa”). Top matches of all four HLA genes were identified in one of the two HLA haplotypes: HLA-A*02:01:01:01, HLA-B*51:01:01, HLA-C*05:01:01:01, and HLA-DRB1*13:01:01 (*SI Appendix, Fig. S7*).

Discussion

In this study, we have demonstrated WGS of single cells using the SISSOR technology in which megabase-size single-stranded DNA fragments from the homologous chromosome pairs of a

single cell are partitioned into multiple compartments for independent amplification and sequencing. With two-strand consensus and long subhaploid fragment assembly, our approach can simultaneously provide higher per base sequencing accuracy and longer haplotypes than other techniques where similar single-cell amplification and sequencing platforms are employed. Unlike the long fragment read method (18), in which errors are reduced by comparing consensus calls from multiple single-stranded libraries from many cells, our SISSOR technology makes use of the consensus of the two complementary strands from only a single cell for error corrections. The long SISSOR fragments also facilitated long-range haplotype assembly with short-read sequencing data acquired from isolated single cells, without the need for extensive cloning (8) and multiplying cells in culture (10). This makes possible high sequencing accuracy using rare single cells or nondividing cells from primary tissues, such as adult neurons (19), where brain function influenced by single-cell genomic diversity and the somatic mutations could heighten the risk of developing certain neuropsychiatric disorders. Using a reference genome, the accuracy of two-strand variant consensus calls was confirmed to be better than 1 error in 100 million bases, which is a major improvement from 2,000 errors in 100 million bases obtained using the double-stranded sequencing library. The accuracy exceeds what is required for detecting the potential genetic variations between single cells (10^{-7}). The ability to determine long haplotypes will also enable applications such as accurate HLA haplotyping for better donor–patient matching for organ transplants. Our current implementation of the SISSOR technology still has limitations, including the lack of integrated sequencing library preparation and scalability for parallel processing and sequencing of multiple single cells. These limitations can be addressed by designing a microfluidic processor with polymer barriers to enable single-chamber multistep processing required for on-chip preparation of encoded sequencing libraries (20) and integrating the water-in-oil droplet-based approach to enable parallel genome sequencing of multiple single cells. We anticipated diverse research and clinical needs will be enabled by highly accurate genome sequencing and haplotyping of single cells.

Materials and Methods

Fabrication of Microfluidic Processors. The PDMS microfluidic devices with domed flow channels at the top layer and 25 μm -thick valve control layer at the bottom were fabricated using soft lithography. Both layers are bonded together and then to a glass slide using standard PDMS techniques (20, 21). The mold for the fluidic channels with multiple heights was fabricated by repeating the soft lithography process. All domed channels (12 μm and 20 μm) were fabricated using photoresist AZ 12XT-20PL-10 (AZ) followed by reflowing to produce the domed structure. The 60 μm -tall rectangle MDA chambers are fabricated using SU8 2050 (MicroChem). The molds were passivated with tridecafluoro-1,1,2,2-tetra-hydrooctyl-1-trichlorosilane (Pfaltz and Bauer, Inc.) before casting PDMS. The PDMS layers were fabricated using Sylgard 184 (Dow Corning). A 5:1 mixture (part A:part B = 5:1) was poured onto the mold for the top fluidic layer. A 20:1 mixture (part A:part B = 20:1) was spin-coated onto the mold for the valve control layer at 1,500 rpm for 45 s in a spin coater (Laurell WS-650-23B). After both PDMS layers were cured on the molds for 25 min at 65 $^{\circ}\text{C}$ in an oven, the fluidic layer was peeled off and access holes were created using a 0.75 mm-diameter biopsy punch (Ted Pella, Inc.). The fluid layer was then aligned and laid onto the thin valve control layer. The two PDMS layers were bonded for 4 h at 65 $^{\circ}\text{C}$ in an oven. The bonded layers were peeled off, and the valve-connecting holes were punched. The surface of the valve layer and a cover glass (75 mm \times 50 mm \times 1.0 mm) were treated with oxygen plasma in a UV-ozone cleaner (Jelight Company, Inc.) for 4 min and then bonded together for 10 h at 65 $^{\circ}\text{C}$ in an oven. A photograph of a working device is shown in *SI Appendix, Fig. S1*.

Partitioning, Amplification, and Sequencing of Single-Cell Genome. On-chip amplification was performed with a modified protocol of the Nextera Phi29 kit. After filling all valve lines with pure filtered water (18 M Ω -cm), all MDA chambers were incubated for 15 min with 0.1% BSA, 35 mM random

hexamers, and 16 μM dNTPs and then purged and dried with airflow. The entire device was then sterilized for 15 min in an UV crosslinker (Longwave UV Crosslinker; UVP). A single cell suspended in PBS buffer was immediately loaded in the capture chamber and then flushed and filled with alkaline lysis solution (ALS) (400 mM KOH, 10 mM DTT, and 1% Tween20) in the mixing chamber. The cell lysate was mixed with the lysis solution in the ring mixer for 10 min using the PDMS peristaltic pump operating at 5 Hz and then loaded in the 24 partition chambers by pushing with air. A neutralization solution (NS) (400 mM HCl, 600 mM Tris-HCl, and 1% Tween20) and Nextera MDA mastermix (1 \times MDA buffer, 84 μM 3' phosphorylated random hexamers, 2 mM dNTPs, 150 μM dUTP, and 0.84 μg of Phi29 DNA polymerase) were loaded into a channel parallel to the partition chambers, and all partitioning valves were closed along the NS line and partitioning chambers. The reaction mix was pushed into the individual MDA chamber, and the device was incubated at 30 $^{\circ}\text{C}$ for 15 h to carry out the MDA reactions. All of the feed lines were flushed with air and washed with TE buffer. After amplification, the MDA products in each chamber were pushed into a pipet tip with 1 \times TE buffer at the individual outlet, and 5 μL was collected. The samples were incubated at 65 $^{\circ}\text{C}$ for 15 min to inactivate Phi29 DNA polymerase. Construction of barcoded sequencing libraries and Illumina

sequencing were performed as described by Peters et al. (18) and Adey et al. (22) and further described in *SI Appendix, SI Methods*.

Mapping Sequencing Reads. Human genome extension hs37d5 (GRCh37 + decoy) from the 1000 Genomes Project was used as the reference for mapping. All reads were mapped with BWA-MEM (23). Paired-end 100-bp reads were mapped as two single-end reads. All reads in the fastq format were aligned to the human genome reference in the fasta format. Alignments were obtained in the SAM/BAM format (24). For BWA-MEM alignment, the "mem -M" command was applied to BWA (version 0.7.10). Distributions of unprocessed sequencing reads of the first five chromosomes from each single cell were binned and displayed in *SI Appendix, Figs. S8–S22*. Direct comparison in the same 1-Mbp region of chromosome 15 in Peters et al. (18) is shown in *SI Appendix, Figs. S23 and S24*.

ACKNOWLEDGMENTS. This work was supported by NIH Grants R01HG007836 (to K.Z. and X.H.) and R01GM114362 (to V. Bafna) and was performed in part at the San Diego Nanotechnology Infrastructure of the University of California, San Diego, a member of the National Nanotechnology Coordinated Infrastructure, which is supported by National Science Foundation Grant ECCS-1542148.

- Ke X, Taylor MS, Cardon LR (2008) Singleton SNPs in the human genome and implications for genome-wide association studies. *Eur J Hum Genet* 16:506–515.
- Collins FS, Brooks LD, Chakravarti A (1998) A DNA polymorphism discovery resource for research on human genetic variation. *Genome Res* 8:1229–1231.
- Fu Y, et al. (2015) Uniform and accurate single-cell sequencing based on emulsion whole-genome amplification. *Proc Natl Acad Sci USA* 112:11923–11928.
- Wheeler DA, et al. (2008) The complete genome of an individual by massively parallel DNA sequencing. *Nature* 452:872–876.
- Wang J, et al. (2008) The diploid genome sequence of an Asian individual. *Nature* 456:60–65.
- Pushkarev D, Neff NF, Quake SR (2009) Single-molecule sequencing of an individual human genome. *Nat Biotechnol* 27:847–850.
- McKernan KJ, et al. (2009) Sequence and structural variation in a human genome uncovered by short-read, massively parallel ligation sequencing using two-base encoding. *Genome Res* 19:1527–1541.
- Lo C, et al. (2013) On the design of clone-based haplotyping. *Genome Biol* 14:R100.
- Fan HC, Wang J, Potanina A, Quake SR (2011) Whole-genome molecular haplotyping of single cells. *Nat Biotechnol* 29:51–57.
- Porubsky D, et al. (2016) Building complete chromosomal haplotypes using single-cell sequencing. *Genome Res*, 10.1101/gr.209841.116.
- Zheng GXY, et al. (2016) Haplotyping germline and cancer genomes with high-throughput linked-read sequencing. *Nat Biotechnol* 34:303–311.
- Baslan T, et al. (2012) Genome-wide copy number analysis of single cells. *Nat Protoc* 7:1024–1041.
- Edge P, Bafna V, Bansal V (2016) HapCUT2: Robust and accurate haplotype assembly for diverse sequencing technologies. *Genome Res* 27:801–812.
- Paez JG, et al. (2004) Genome coverage and sequence fidelity of phi29 polymerase-based multiple strand displacement whole genome amplification. *Nucleic Acids Res* 32:e71.
- Chen L, Liu P, Evans TC, Jr, Ettwiller LM (2017) DNA damage is a pervasive cause of sequencing errors, directly confounding variant identification. *Science* 355:752–756.
- Ball MP, et al. (2012) A public resource facilitating clinical use of genomes. *Proc Natl Acad Sci USA* 109:11920–11927.
- Consortium TEP; ENCODE Project Consortium (2012) An integrated encyclopedia of DNA elements in the human genome. *Nature* 489:57–74.
- Peters BA, et al. (2012) Accurate whole-genome sequencing and haplotyping from 10 to 20 human cells. *Nature* 487:190–195.
- McConnell MJ, et al. (2017) Intersection of diverse neuronal genomes and neuropsychiatric disease: The Brain Somatic Mosaicism Network. *Science* 356:eaal1641.
- Lee HS, Chu WK, Zhang K, Huang X (2013) Microfluidic devices with permeable polymer barriers for capture and transport of biomolecules and cells. *Lab Chip* 13:3389–3397.
- Unger MA, Chou HP, Thorsen T, Scherer A, Quake SR (2000) Monolithic micro-fabricated valves and pumps by multilayer soft lithography. *Science* 288:113–116.
- Adey A, et al. (2010) Rapid, low-input, low-bias construction of shotgun fragment libraries by high-density in vitro transposition. *Genome Biol* 11:R119.
- Li H (2013) Aligning sequence reads, clone sequences and assembly contigs with BWA-MEM. arXiv:1303.3997.
- Li H, et al.; 1000 Genome Project Data Processing Subgroup (2009) The sequence alignment/map format and SAMtools. *Bioinformatics* 25:2078–2079.