



2015-06-01

Will Ultrasound Performed with the Rich-Mar AutoSound™ Be as Effective at Increasing Tissue Temperature as Ultrasound Performed with a Traditional Machine?

Heather Diane Black
Brigham Young University - Provo

Follow this and additional works at: <https://scholarsarchive.byu.edu/etd>

 Part of the [Exercise Science Commons](#)

BYU ScholarsArchive Citation

Black, Heather Diane, "Will Ultrasound Performed with the Rich-Mar AutoSound™ Be as Effective at Increasing Tissue Temperature as Ultrasound Performed with a Traditional Machine?" (2015). *All Theses and Dissertations*. 5884.
<https://scholarsarchive.byu.edu/etd/5884>

This Thesis is brought to you for free and open access by BYU ScholarsArchive. It has been accepted for inclusion in All Theses and Dissertations by an authorized administrator of BYU ScholarsArchive. For more information, please contact scholarsarchive@byu.edu, ellen_amatangelo@byu.edu.

Will Ultrasound Performed with the Rich-Mar AutoSound™ Be as Effective
at Increasing Tissue Temperature as Ultrasound Performed
with a Traditional Machine?

Heather Diane Black

A thesis submitted to the faculty of
Brigham Young University
in partial fulfillment of the requirements for the degree of
Master of Science

David O. Draper, Chair
Ulrike H. Mitchell
Dennis L. Eggett

Department of Exercise Sciences

Brigham Young University

June 2015

Copyright © 2015 Heather Diane Black

All Rights Reserved

ABSTRACT

Will Ultrasound Performed with the Rich-Mar AutoSound™ Be as Effective at Increasing Tissue Temperature as Ultrasound Performed with a Traditional Machine?

Heather Diane Black
Department of Exercise Sciences, BYU
Master of Science

STUDY DESIGN: Randomized crossover experiment. **OBJECTIVE:** To determine whether the Rich-Mar AutoSound™ would be as effective as traditional ultrasound at increasing the temperature of the triceps surae muscle during a 10-min, 1 MHz, 1.0 W/cm² ultrasound treatment. **BACKGROUND:** The AutoSound™ is a hands-free ultrasound device that is strapped on the body and left for the duration of the ultrasound treatment. It requires no clinician during the actual ultrasound treatment, thus freeing the clinician to perform other tasks and reducing clinician error during treatments. **METHODS:** 16 healthy subjects (6 males, 10 females, age = 22 ± 1.6 yrs, height = 173.2 ± 8.4 cm, weight = 72.5 ± 11.3 kg, triceps surae subcutaneous fat thickness = 0.85 ± 0.37 cm) received a 10-min, 1 MHz, 1.0 W/cm² ultrasound treatment over their left triceps surae muscle with both the AutoSound™ and traditional ultrasound (via the TheraHammer™) with 24 hours between treatments. Temperatures were measured every 30 seconds during the ultrasound treatments by way of a thermistor, approximately 2.25 cm deep in the triceps surae. **RESULTS:** The AutoSound™ was not effective at increasing the temperature of the triceps surae muscle, as temperature decreased 0.16°C during treatment (p = 0.334). On average, the AutoSound™ caused intramuscular temperature to decrease at a rate of 0.016 ± 0.001°C per min. Traditional ultrasound performed using the TheraHammer™ had a total temperature increase of 0.41°C. Rate of temperature increase during traditional ultrasound was 0.025 ± 0.003°C per min (p < 0.0001). **CONCLUSION:** The AutoSound™ is not as effective at increasing muscle temperature as traditional ultrasound during a 10-min, 1 MHz, 1.0 W/cm² treatment. However, neither the AutoSound™ nor traditional ultrasound was very effective at increasing the temperature of the triceps surae muscle during the treatment time.

Keywords: AutoSound™, ultrasound, intramuscular temperature changes

ACKNOWLEDGEMENTS

I would sincerely like to thank my committee chair, Dr. Draper, for his support and encouragement through this whole process. I would have been completely lost without his help and appreciate the lengths he went to both with my thesis and in making me feel welcome at BYU. I also want to thank my committee members for all of the effort they put in during these years on my behalf. Dr. Mitchell for her never-ending support and her thoroughness in critiquing my work and making me better, and Dr. Eggett for his help with the statistical analysis and his patience in answering my statistically-related questions. I have appreciated their guidance and support and the knowledge that they truly care about their students. I also want to thank my parents for the support and encouragement they have given me my whole life. It is by their example that I have learned to work hard and persevere and to be proud of the work that I do. Lastly I want to thank my best friend and fiancé Thomas James. He has been the rock for me to lean against this past year, and his never-failing optimism and belief in me has made me better in every way. I am excited to finish one chapter of my life and begin another with him. I really appreciate the opportunity I have had during graduate school to learn and to grow and am truly grateful for everyone that has helped me along the way.

TABLE OF CONTENTS

Title Page	i
Abstract	ii
Acknowledgements	iii
Table of Contents	iv
List of Tables	v
List of Figures	vi
Introduction	1
Methods	3
Participants	3
Instrumentation	3
Procedures	4
Catheter and Thermistor Insertion	4
Ultrasound Treatment Area	5
Manual Ultrasound Treatment	5
AutoSound™ Ultrasound Treatment	5
Thermistor Removal	6
Statistical Analysis	6
Results	6
Discussion	7
Limitations and Future Research	12
Conclusion	12
References	13

LIST OF TABLES

TABLE	PAGE
1. Summary of baseline, final, total, and rate of temperature change (mean \pm standard error).....	16

LIST OF FIGURES

FIGURE	PAGE
1. Temperature probe insertion on the medial side of the left triceps surae	17
2. Verification of probe depth insertion via Doppler ultrasound imaging	18
3. Ultrasound performed with the AutoSound™ machine	19

INTRODUCTION

Therapeutic ultrasound is one of the most common deep heating modalities used by physical therapists, athletic trainers, and occupational therapists.¹ The thermal effects of ultrasound are: treating soft tissue injuries² and muscle spasm,³ restoring range of motion,⁴ increasing collagen extensibility,⁵ aiding in collagen alignment^{6,7} and increasing wound strength.⁷ The nonthermal effects of ultrasound include: increasing histamine release,¹ increasing phagocytosis,^{1,8} increasing protein synthesis,⁹ enhancing tissue regeneration^{8,10} and wound healing,¹¹ and increasing fibroblasts and vascular regeneration.¹⁰⁻¹² Therapeutic ultrasound uses high frequency, inaudible, acoustic vibrations to produce these thermal and nonthermal physiological effects. Unfortunately, despite its common use, therapeutic ultrasound is often misunderstood and misused.^{13, 10, 12} However, when used properly, it is an effective treatment method that can be applied to both normal and damaged tissue.^{4,14-17}

Traditional ultrasound treatments are prone to clinician error (treating too large a surface area,¹³ moving the soundhead faster than the recommended speed,¹⁸ etc.), labor intensive and time consuming, requiring a clinician to manually move the ultrasound transducer over the target tissue, leaving the clinician occupied and unable to complete other tasks. Rich-Mar (Chattanooga, TN) addressed these problems by developing the AutoSound™, a hands-free ultrasound alternative. The AutoSound™ works by activating and deactivating four rectangular transducer crystals that lie side-by-side.¹⁹ The first crystal turns on and then quickly turns off when the second crystal turns on. This process repeats down to crystal four, and then starts at crystal one again. The activation and deactivation of the crystals is equivalent to a clinician manually moving the ultrasound transducer at a speed of 4 cm/sec,²⁰ the recommended speed of traditional soundhead movement.²¹⁻²³ The firing pattern of the crystals is equivalent to manually

moving the ultrasound transducer from one part of the treatment area to the other, picking the transducer up, and placing it back at the starting point.²⁴ These crystals are housed in one unit that can be strapped on the body and left for the duration of the treatment.

The AutoSound™ could be a tremendous clinical asset, significantly adding to the time efficiency of the clinician if the machine works. Multiple studies,²⁴⁻²⁶ have compared the AutoSound™ against traditional ultrasound in its ability to heat human muscles, and all have found that traditional ultrasound produced significantly greater temperature increases than the AutoSound™. Upon further examination of these comparison studies,²⁴⁻²⁶ we discovered that each of the three used the 3 MHz frequency. Ultrasound delivered at a 3 MHz frequency is absorbed superficially in the tissues 1–2 cm deep, but may reach all the way to 3 cm deep,^{27,28} whereas ultrasound delivered at 1 MHz is absorbed in deeper tissues 2–5 cm deep.²⁷

The intensity used in these studies is important as well. Intensity is the rate at which ultrasound waves are being delivered to target tissues per unit area of the transducer surface (expressed as W/cm²).²⁹ The lower the intensity, the longer the treatment duration needs to be in order to achieve the desired results.²⁷ Two of the previous studies^{25,26} on the AutoSound™ used an intensity of 1.5 W/cm² for 10 min. The other²⁴ used 1.0 W/cm² for 8 min, even though treatments at a lower intensity should be longer duration to produce the desired results. The purpose of this study was to compare intramuscular temperature changes produced by a 10-min ultrasound treatment via the AutoSound™ and a traditional ultrasound treatment at a frequency of 1 MHz and an intensity of 1.0 W/cm².

METHODS

Participants

We recruited 16 healthy subjects for this study (age = 22 ± 1.6 y, triceps surae subcutaneous fat thickness = 0.85 ± 0.37 cm, 6 males, 10 females; weight = 72.5 ± 11.3 kg; height = 173.2 ± 8.4 cm). Subjects were screened for exclusion criteria during their signing of the consent form. Exclusion criteria were: a lower extremity injury within the last two months, a lower leg infection, open wound, rash, swelling, ecchymosis, decreased circulation, decreased sensation in the area being treated or thrombophlebitis. Participants refrained from exercise 2 h prior to each lab visit. All subjects provided written consent before their participation in the study. The study was approved by the University's Institutional Review Board before subject recruitment began.

Instrumentation

Traditional ultrasound was produced via the TheraSound Evo™ (Rich-Mar, Chattanooga, TN) delivered at a frequency of 1MHz. All Rich-Mar ultrasound machines use a beam nonuniformity ratio of 5.5:1 or less, and have an effective radiating area as close to the size of the soundhead as possible.¹⁹ The traditional ultrasound was performed using the TheraHammer™ (Rich-Mar, Chattanooga, TN) which houses a lead zirconate titanate crystal that is 2 cm². Hands-free ultrasound was performed using the TheraSound Evo™ with the AutoSound™ attachment. The four crystals of the AutoSound™ are 1.5 cm by 2.5 cm each with 2 mm of dead space between each crystal. The treatment area of the AutoSound™ is approximately 14 cm².

Temperature was measured using the ISO-Thermex (Columbus Instruments International, Inc., Columbus, OH) program. Temperature readings were received from an IT-21 thermistor (Physitemp Instruments Inc., Clifton, NJ). The thermistor was inserted via a 20-gauge catheter

(BD Medical, Franklin Lakes, NJ). The depth of the inserted thermistor as well as adipose thickness (adipose levels were taken from 3 locations on each subject: one directly above the thermistor, one on the far left of the frozen image and one on the far right. All measurements were marked from the bottom of the skin to the top of the fascia surrounding the triceps surae muscle, with all measurements averaged together) was measured using Doppler ultrasound imaging (model: LogiQ 55e, General Electric Company, Fairfield, CT). Imaging ultrasound was produced using the 12 L soundhead and a 12 MHz frequency.

Procedures

A randomized cross-over experiment was performed. Participants reported to the lab twice, with at least 24 h between visits. All participants were screened for contraindications via consent form, and those who were still eligible after the screening process reviewed and signed an Institutional Review Board approved consent form. Once subjects were officially enrolled in the study, they were randomly assigned by drawing a piece of paper out of an opaque cup to receive ultrasound treatment first via the AutoSound™ or by using traditional ultrasound.

Catheter and Thermistor Insertion

A single thermistor was inserted via catheter into the medial side of the subject's left lower leg, an average depth of 2.25 ± 0.52 cm in the tissue (see FIGURE 1). Patients laid prone on the treatment table during catheter insertion (and for the remaining time of the treatment) with their left lower leg exposed. The area of greatest girth on the patient's triceps surae was visualized. A T-square was used to measure 2.25 cm anterior on the medial triceps surae and a green dot was marked on the skin where the catheter would be inserted. The insertion site was cleaned with an iodine swab and allowed to air dry before the catheter was inserted. The catheter was horizontally inserted into the medial triceps surae muscle over the previously marked green

spot. A thermistor was fed through the catheter, and the catheter was removed, leaving the thermistor in place. Following the methods of another study²⁴ we used only one depth for the thermistors at 2.25 cm deep. Ultrasound treatment at 1 MHz ideally targets tissues 2–5 cm deep.²⁷ Thermistor insertion depth was verified using Doppler ultrasound imaging (see FIGURE 2).

Ultrasound Treatment Area

Ultrasound treatment area was centered over the end of the thermistor for each subject. The treatment area of traditional ultrasound was marked using a template two times the size of the ultrasound head (approximately 4 cm²). Treatments performed via the AutoSound™ covered approximately 14 cm².

Manual Ultrasound Treatment

The ultrasound treatment using the manual technique was administered within the previously marked spot on the back of the triceps surae for 10 min. Treatments were performed with a 2 cm² transducer at a frequency of 1 MHz and an intensity of 1.0 W/cm². Tissue temperature readings were recorded for each patient at baseline (once temperature had stabilized to the point that there was no more than 0.2°C change every 30 sec) and every 30 sec for the duration of the treatment. Ultrasound gel was used as the coupling medium in all traditional treatments.

AutoSound™ Ultrasound Treatment

Treatments performed with the AutoSound™ (see FIGURE 3) were performed on the same leg as the manual ultrasound treatment, once again over the area of greatest girth on the medial triceps surae. Settings of 1 MHz and 1.0 W/cm² for 10 min were used. The AutoSound™ was secured in place with 1-inch Powerflex tape. Ultrasound treatments were started after tissue

temperature had stabilized to the point that there was no more than 0.1°C change every 30 sec. Intramuscular temperatures were recorded every 30 sec throughout the treatment session using the ISO-Thermex. A 1 cm thick gel pad (designed specifically for the AutoSound™) was used as the coupling medium during all AutoSound™ treatments.

Thermistor Removal

At the conclusion of each treatment, the thermistor was removed from the subject's triceps surae and a bandage was placed over the area for protection. The thermistors and catheters were sterilized using an Anprolene Gas Sterilizer (Model: AN74i, Andersen Products, Inc., Haw River, NC).

Statistical Analysis

A 2x2 repeated measures ANOVA was used to determine interactions among the beginning and ending temperatures of each ultrasound unit. A hierarchical linear model was used to determine the rate of temperature change caused by each machine. In this model a regression line was fit to the slope of temperature change for each individual. Individual slopes were then averaged together for an overall slope of the population. SAS 9.3 (2010) was used for all statistical analysis, and alpha was set at $p < 0.05$.

RESULTS

The 2x2 repeated measures ANOVA on temperature showed a statistically significant interaction between instruments and time ($F = 23.72$ ($p = .0002$)). On average, traditional ultrasound temperatures ranged from a starting temperature of $35.67^{\circ}\text{C} \pm$ standard error of 0.24°C to an ending temperature of $36.08^{\circ}\text{C} \pm 0.24^{\circ}\text{C}$. Mean tissue temperature before ultrasound performed with the AutoSound™ was $35.88^{\circ}\text{C} \pm 0.24^{\circ}\text{C}$ and ending temperature was $35.73^{\circ}\text{C} \pm 0.24^{\circ}\text{C}$. Traditional ultrasound mean changes between beginning and ending

temperatures were statistically significant, with traditional ultrasound increasing tissue temperature $0.41^{\circ}\text{C} \pm 0.09$ ($p = .0016$). There was no statistically significant change from beginning to ending temperature (tissue temp went down 0.16°C) with ultrasound performed with the AutoSound™ ($p = 0.33$).

The hierarchical linear model revealed a statistically significant difference in the slopes between traditional ultrasound and the AutoSound™ $F = 124.17$ ($p < .0001$) with regard to the rate of heating. On average, traditional ultrasound increased tissue temperature $0.025^{\circ}\text{C}/\text{min} \pm 0.003^{\circ}\text{C}$ ($p < .0001$). The AutoSound™ actually lowered tissue temperature $0.016^{\circ}\text{C} \pm 0.001^{\circ}\text{C}/\text{min}$ ($p = 0.95$; see TABLE 1).

DISCUSSION

We compared the heating of the AutoSound™ with traditional ultrasound delivered by the TheraSound Evo™ using the TheraHammer™. We discovered that the AutoSound™ at 1 MHz did not raise the tissue temperature during the 10-min treatment. These findings support previous research²⁴⁻²⁶ that the AutoSound™ does not heat as well as traditional ultrasound. Three studies have compared the heating of the AutoSound™ with traditional ultrasound at a frequency of 3 MHz,²⁴⁻²⁶ though ours is the first to test the AutoSound™ at 1 MHz. McCutchan et al.²⁴ used the following parameters for their study: 3 MHz, $1.0 \text{ W}/\text{cm}^2$, 8 min, assessing the tissue temperature at a depth of 1 cm. They found a 1.8°C increase in tissue temperature when the AutoSound™ was used and a 3.2°C increase when the Omnisound™ (Accelerated Care Plus, Reno, NV) was used. Like McCutchan et al.,²⁴ we used an intensity of $1.0 \text{ W}/\text{cm}^2$, but a longer treatment time of 10 min. In both cases traditional ultrasound produced a significantly higher increase in tissue temperature when compared with the AutoSound™.

The following parameters were used in the Gulick²⁵ study: 3 MHz, 1.5 W/cm², 10 min, tissue temperature probes 1 and 2 cm deep. The AutoSound™ increased the tissue temperature 5.1°C at 1 cm deep, and 1.5°C at 2 cm deep. The Omnisound™ increased the tissue temperature 6.7°C at 1 cm and 4.0°C at 2 cm. Traditional ultrasound once again produced a significantly greater increase in tissue temperature when compared with the AutoSound™. Though our settings varied from Gulick in every other way, we also used a 10-min treatment time.

Fincher et al.²⁶ performed ultrasound using the AutoSound™ at 3 MHz and 1.5 W/cm² for 10 min and traditional ultrasound via the 5 cm² TheraHammer™ transducer on the AutoSound™ 7.6 Combo unit at a depth of 2.5 cm. The AutoSound™ increased temperature 2.05°C, while traditional ultrasound increased tissue temperature 4.53°C. Again traditional ultrasound produced significantly higher temperature increases than the AutoSound™. We, like Fincher et al.,²⁶ used the same ultrasound machine with different attachments for all ultrasound treatments, but used the 2 cm² transducer for the traditional treatment instead of the 5 cm² transducer that was used in this study. The 2 cm² transducer or the difference in frequency may explain why Fincher et al.²⁶ received a 4.53°C change and a 2.05°C change with traditional ultrasound and the AutoSound™, respectively, and we saw very little change in temperature.

The AutoSound™ did not increase tissue temperature to the same degree as traditional ultrasound in any of these cases.²⁴⁻²⁶ Research has found that heating varies from manufacturer to manufacturer.³⁰⁻³² The Omnisound™ was used in two of these studies^{24,25} and may heat at a different rate than the TheraSound Evo™ as it seems to increase tissue temperature more than any other ultrasound machine with which it has been compared.^{33, 34} To eliminate variability between manufacturers, we, like Fincher et al.,²⁶ compared 2 devices manufactured by the same company (Rich-Mar).

This is the first study performed on the AutoSound™ at a 1 MHz frequency. Even though there is variability between manufacturers, studies with similar parameters can help give an estimate of temperature changes that would be expected. Demchak et al.³⁰ found at a depth of 3 cm, a 1 MHz, 1.2 W/cm², 10-min treatment from the OmniSound 3000C increased tissue temperature 0.3°C per minute. The same parameters at 1.5 W/cm² increased muscle temperature at a rate of 0.4°C per minute.³⁵ At 2.5 cm, a 1 MHz, 1.5 W/cm² treatment increased temperature 0.26°C per minute.³⁶ The same parameters at 1.0 W/cm² increased temperature 0.16 ± 0.072°C per minute.²⁷ Thus, if our study followed the heating rate found in other studies with similar parameters, intramuscular temperature should have increased anywhere from 0.16–0.40°C. Instead our heating rate with traditional ultrasound was 0.025 ± 0.003°C per minute and -0.016 ± 0.001°C with the AutoSound™.

In our opinion the following are reasons why the AutoSound™ did not raise the tissue temperature: First, there is a slight time lag between the firing of each successive crystal. This means that there is not always a crystal on, which could lead to a decrease in heating. There is also a slight amount of space (2 mm) between each of the four crystals in the AutoSound™. This slight space between each crystal means that there is “dead space” where no heating occurs in the ultrasound unit. This may effect target tissue temperature change. The time delay from one crystal to the next and the fact that there is no heating under the dead space between adjacent crystals¹ could be a reason the AutoSound™ does not appear to heat the tissue to the same degree as traditional ultrasound.

Second, the gel pad used during AutoSound™ application may be too thick. Ultrasound gel has been the coupling medium used during all traditional ultrasound treatments in the studies where traditional ultrasound was compared to the AutoSound™.²⁴⁻²⁶ Studies have shown that

ultrasound gel (such as used during traditional ultrasound) is the most effective form of coupling medium at increasing tissue temperature when compared to 1 cm or 2 cm gel pads.^{37,38} The AutoSound™, however, uses a gel pad that is 1 cm thick. This may impair the ultrasound unit's ability to effectively deliver sound waves into the target tissue. Recent research³⁹ has shown that the Gel Shot™ (a 2–3 mm thick gel pad) is more effective than ultrasound gel when used at 1 MHz. Therefore it is possible that a thinner gel pad could have been more effective and aid in increasing tissue temperature, but the gel pad currently used with the AutoSound™ may be too thick to see any positive effects.

The third reason the AutoSound™ may not be effective in tissue heating has to do with the activation sequence and arrangement of the 4 crystals. A traditional ultrasound transducer uses one crystal. Ultrasound only produces significant heating when an area 2 times the size of the soundhead is used.^{1,18,33,40} Chudleigh et al.⁴¹ found that at 3 cm, a 10-min, 1 MHz, 1.5 W/cm² ultrasound treatment resulted in a 3.5°C increase in temperature when an area 2 times the size of the soundhead was treated. However, an area 6 times the size of the soundhead increased the temperature only 0.57°C. Thus a larger treatment area leads to a decrease in the amount of heating that takes place. During traditional ultrasound an area the size of 2 crystals is heated. Although the AutoSound™ uses 4 crystals that lie side by side, this only mimics one crystal moving as only one crystal is activated at a time. This means that the AutoSound™ is technically covering an area 4 times the size of a soundhead (treating only 25% of the surface area at a time), instead of 2 times the size of the soundhead (treating 50% of the surface area). Additionally, by activating the first crystal every time after crystal 4 has turned on and off, the AutoSound™ is mimicking picking the soundhead up and placing it back at the starting position. Though contact with the skin is maintained during AutoSound™ treatments, mimicking this pattern with

traditional ultrasound would lead to loss of contact with the skin which would lead to a decrease in heating because the sound waves cannot be transduced into the tissue at this point. We suggest that the manufacturers of the AutoSound™ consider placing 2 large crystals (possibly 10 cm²) side by side in a new version of the AutoSound™. Most likely, this would produce higher temperatures.

A temperature increase of 1°C is considered mild heating and is used for increasing metabolism and reducing mild inflammation. A 2–3°C increase is considered moderate heating and is indicated for increasing blood flow and reducing pain and muscle spasm. A 4°C increase is considered to be vigorous heating and is used to increase the extensibility of collagen fibers.¹ According to this, Gulick²⁵ received vigorous heating (5°C increase at 1 cm) and mild heating (1.5°C increase at 2 cm) when the AutoSound™ was used at a frequency of 3 MHz. McCutchan et al.²⁴ and Fincher et al.²⁶ produced moderate heating (1.8°C at 1 cm and 2.05°C at 2.5 cm, respectively) when the AutoSound™ was used, again, at 3 MHz. Heating may have occurred in these studies²⁴⁻²⁶ and not in ours due to the use of 1 MHz. 1 MHz ultrasound heats at 1/3 the peak temperature as 3 MHz.²⁷ This is due to the crystal deforming at 1/3 the rate as a crystal at 3 MHz. Another reason might be that the beam diverges (spreads out) the 1 MHz frequency, whereas the beam is collimated at the 3 MHz frequency.⁴² This might focus more energy on the temperature probe when 3 MHz is used and not increase tissue temperature at 1 MHz.

At a 3 MHz frequency²⁴⁻²⁶ the AutoSound™ may be clinically beneficial as it produces moderate¹ heating. Most clinical practices target superficial tissues, so the AutoSound™ will produce moderate heating in the desired area, as well as free the clinician to perform other tasks. However, at 1 MHz and 1.0 W/cm², the AutoSound™ is not beneficial for clinicians or their patients in heating their tissue. At 1 MHz the AutoSound™ did not produce moderate or even

mild heating. A 20-min hot-pack treatment can raise tissue temperature 3.6°C at 1 cm and 0.8°C at 3 cm.⁴³ Thus the AutoSound™ at 1 MHz is no better than a hot-pack treatment at increasing muscle temperature. However, a hot-pack heats a much larger area than the AutoSound™ making the hot-pack the treatment of choice when targeting deeper tissues.

Limitations and Future Research

Our study had limitations. We used healthy subjects from 18–25 years of age to examine tissue temperatures in nondamaged tissue. We assume that tissue temperature changes would be similar in an injured population over damaged tissue. Our results are also limited to the use of a 2 cm² soundhead, a frequency of 1 MHz, and an intensity of 1.0 W/cm².

We suggest future research should be conducted on the AutoSound™ at 1 MHz and at a higher intensity than 1.0 W/cm².

CONCLUSION

We successfully measured intramuscular temperature changes during ultrasound treatment with a traditional and a hands-free device. At a depth of 2.25 cm, a 10-min, 1 MHz, 1.0 W/cm² ultrasound treatment did not produce desired heating with either machine. At 1 MHz, the AutoSound™ failed to increase the temperature of the triceps surae muscle, and the TheraHammer™ only minimally increased temperature. We suggest an alteration to the AutoSound™ to where only two larger crystals are used so an area twice the size of the soundhead is treated. We also suggest employing the use of a thinner gel pad during AutoSound™ treatment.

REFERENCES

1. Knight KL, Draper DO. *Therapeutic Modalities The Art and Science*. 2nd ed. Philadelphia, PA: Lippincott Williams & Wilkins 2013.
2. Binder A. Is therapeutic ultrasound effective in treating soft tissue lesions? *British Medicine Journal*. 1985(290):512-514.
3. Draper DO, Mahaffey C, Kaiser D, Eggett D, Jarmin J. Thermal ultrasound decreases tissue stiffness of trigger points in upper trapezius muscles. *Physiotherapy Theory & Practice*. 2010;26(3):167-172.
4. Draper DO. Ultrasound and joint mobilizations for achieving normal wrist range of motion after injury or surgery: A case series. *Journal of Athletic Training*. 2010;45(5):486-491.
5. Rose S, Draper DO, Schulthies SS, Durrant E. The stretching window part two: rate of thermal decay in deep muscle following 1-MHz ultrasound. *Journal of Athletic Training*. Apr 1996;31(2):139-143.
6. da Cunha A, Parizotto NA, Vidal BC. The effect of therapeutic ultrasound on repair of the achilles tendon (tendo calcaneus) of the rat. *Ultrasound in Medicine & Biology*. Dec 2001;27(12):1691-1696.
7. Byl NN, McKenzie A, Wong T, West J, Hunt TK. Incisional wound healing: a controlled study of low and high dose ultrasound. *The Journal of Orthopaedic and Sports Physical Therapy*. Nov 1993;18(5):619-628.
8. Johns LD. Nonthermal effects of therapeutic ultrasound: the frequency resonance hypothesis. *Journal of Athletic Training*. Jul 2002;37(3):293-299.
9. Stewart H, Stratmeyer ME. An Overview Of Ultrasound: Theory, Measurement, Medical Applications, and Biological Effects. *Department of Health, Education and Welfare*. Washington, DC. 1982;82:8190.
10. Young SR, Dyson M. The effect of therapeutic ultrasound on angiogenesis. *Ultrasound in Medicine & Biology*. 1990;16(3):261-269.
11. Doan N, Reher P, Meghji S, Harris M. In vitro effects of therapeutic ultrasound on cell proliferation, protein synthesis, and cytokine production by human fibroblasts, osteoblasts, and monocytes. *Journal of Oral and Maxillofacial Surgery*. Apr 1999;57(4):409-419.
12. Hogan RD, Burke KM, Franklin TD. The effect of ultrasound on microvascular hemodynamics in skeletal muscle: effects during ischemia. *Microvascular Research*. May 1982;23(3):370-379.
13. Draper DO. Facts and misfits in ultrasound therapy: steps to improve your treatment outcomes. *European Journal of Physical and Rehabilitation Medicine*. Apr 2014;50(2):209-216.
14. Dyson M, Luke DA. Induction of mast cell degranulation in skin by ultrasound. *IEEE Transactions on Ultrasonics, Ferroelectrics, and Frequency Control*. 1986;33(2):194-201.
15. Warden SJ, Fuchs RK, Kessler CK, Avin KG, Cardinal RE, Stewart RL. Ultrasound produced by a conventional therapeutic ultrasound unit accelerates fracture repair. *Physical Therapy*. Aug 2006;86(8):1118-1127.
16. Warden SJ, Avin KG, Beck EM, DeWolf ME, Hagemeyer MA, Martin KM. Low-intensity pulsed ultrasound accelerates and a nonsteroidal anti-inflammatory drug delays

- knee ligament healing. *The American Journal of Sports Medicine*. Jul 2006;34(7):1094-1102.
17. Stein H, Rosen N, Lerner A, Kaufman H. Minimally invasive surgical techniques for the reconstruction of calcaneal fractures. *Orthopedics*. Oct 2003;26(10):1053-1056.
 18. Draper DO. Ten mistakes commonly made with ultrasound use: current research sheds light on myths. *Athletic Training and Sports Health Care*. 1996;2:95-107.
 19. Rich-Mar. Richmar Autosound 9.6 operation handbook and manual.
 20. RichMar W. AutoSound superior ultrasound. 2011.
 21. Denegar CR. *Therapeutic Modalities for Athletic Injuries*. Champaign, IL: Human Kinetics; 2000.
 22. Draper DO, Prentice WE. *Therapeutic Modalities for Allied Health Professionals*. New York, NY: McGraw-Hill Health Professionals Division; 1998.
 23. Michlovitz S. *Thermal Agents in Rehabilitation*. Philadelphia, PA: F.A. Davis Company; 1996.
 24. McCutchan E, Demchak TJ, Brucker JB. A comparison of the heating efficacy of the Autosound™ with traditional ultrasound methods. *Athletic Training and Sports Health Care*. 2012;4(2):73-78.
 25. Gulick DT. Comparison of tissue heating between manual and hands-free ultrasound techniques. *Physiotherapy Theory & Practice*. 2010;26(2):100-106.
 26. Fincher AL, Trowbridge CA, Ricard MD. A comparison of intramuscular temperature increases and uniformity of heating produced by hands-free AutoSound and manual therapeutic ultrasound techniques. *Journal of Athletic Training*. 2007;42(Supplement):S-41.
 27. Draper DO, Castel J, Castel D. Rate of temperature increase in human muscle during 1 MHz and 3 MHz continuous ultrasound. *Journal of Orthopaedic & Sports Physical Therapy*. 1995/10/01 1995;22(4):142-150.
 28. Franson J, Draper DO, Rigby J, Johnson AW, Mitchell UH. Three MHz ultrasound vigorously heats tissues at a 3 cm depth. *Athletic Training and Sports Health Care*. 2014;6(6):267-272.
 29. Speed CA. Therapeutic ultrasound in soft tissue lesions. *Rheumatology (Oxford)*. Dec 2001;40(12):1331-1336.
 30. Demchak TJ, Straub SJ, Johns LD. Ultrasound heating is curvilinear in nature and varies between transducers from the same manufacturer. *Journal of Sport Rehabilitation*. 2007;16(2):122-130.
 31. Johns LD, Straub SJ, Howard SM. Variability in effective radiating area and output power of new ultrasound transducers at 3 MHz. *Journal of Athletic Training*. Jan-Mar 2007 2007;42(1):22-28.
 32. Johns LD, Straub SJ, Howard SM. Analysis of effective radiating area, power, intensity, and field characteristics of ultrasound transducers. *Archives of Physical Medicine and Rehabilitation*. Jan 2007;88(1):124-129.
 33. Holcomb WR, Joyce CJ. A comparison of temperature increases produced by 2 commonly used ultrasound units. *Journal of Athletic Training*. Jan-Mar 2003 2003;38(1):24-27.
 34. Merrick MA, Bernard KD, Devor ST, Williams JM. Identical 3-MHz ultrasound treatments with different devices produce different intramuscular temperatures. *Journal of Orthopaedic & Sports Physical Therapy*. 2003/07/01 2003;33(7):379-385.

35. Demchak TJ, Stone MB. Effectiveness of clinical ultrasound parameters on changing intramuscular temperature. *Journal of Sport Rehabilitation*. 2008;17(3):220-229.
36. Miller MG, Longoria JR, Cheatham CC, Baker RJ, Michael TJ. Intramuscular temperature differences between the mid-point and peripheral effective radiating area with ultrasound. *Journal of Sports Science & Medicine*. 2008 2008;7(2):286-291.
37. Bishop S, Draper DO, Knight KL, Feland JB, Eggett D. Human tissue-temperature rise during ultrasound treatments with the Aquaflex gel pad. *Journal of Athletic Training*. Apr-Jun 2004 2004;39(2):126-131.
38. Draper DO, Edvalson CG, Knight KL, Eggett D, Shurtz J. Temperature increases in the human achilles tendon during ultrasound treatments with commercial ultrasound gel and full-thickness and half-thickness gel pads. *Journal of Athletic Training*. 2010;45(4):333-337.
39. Draper DO, Rigby JH, Wells AM. The Gel Shot: An improvement in ultrasound coupling media. *Athletic Training and Sports Health Care (under review)*. 2014.
40. Myrer JW, Measom GJ, Fellingham GW. Intramuscular temperature rises with topical analgesics used as coupling agents during therapeutic ultrasound. *Journal of Athletic Training*. Jan-Mar 2001;36(1):20-25.
41. Chudleigh D, Schulthies S, Draper D, Myrer J. Muscle temperature rise during 1MHz ultrasound treatments of two and six times the effective radiating areas of the transducer. *Journal of Athletic Training*. 1998;33:s11.
42. Starkey C. *Therapeutic Modalities* 4th ed. Philadelphia: FA Davis Company; 2013.
43. Draper DO, Harris ST, Schulthies S, Durrant E, Knight KL, Ricard M. Hot-pack and 1-MHz ultrasound treatments have an additive effect on muscle temperature increase. *Journal of Athletic Training*. Jan 1998;33(1):21-24.

TABLE 1 Summary of baseline, final, and total temperature change, as well as rate of temperature change (mean \pm standard error)

Mode of Change Treatment	Baseline Temperature ($^{\circ}$ C)	Final Temperature ($^{\circ}$ C)	Total Change ($^{\circ}$ C)	Rate of Temperature Change ($^{\circ}$ C/min)
Traditional	35.67 \pm 0.24	36.08 \pm 0.24	0.41	0.025 \pm 0.003
AutoSound TM	35.88 \pm 0.24	35.73 \pm 0.24	-0.16	-0.016 \pm 0.001



FIGURE 1 Temperature probe insertion on the medial side of the left triceps surae



FIGURE 2 Verification of probe depth insertion via Doppler ultrasound imaging



FIGURE 3 Ultrasound performed with the AutoSound™ machine